HPLC 19 - Agilent Lab Advisor

Agilent Lab Advisor

Service & Diagnostics

Advanced Version

Tasks

Spectral Scan

HPLC 19 » HPLC 19 » G7114A:DEACX26660

Lab Advisor

System Overview

Configuration

Apps

Firmware Update

Logs & Results

HPLC 19

Service & Diagnostics

Instrument Control

EMFs

System Report

Help

Context Help (F1)


Help Topics

What's New?

Spectral Scan - 1260 VWD [DEACX26660]

UV Lamp On | Blank Scan

Sample Scan | from 200 | to 300 | step 1 | Export Data



more... | Print Result | Back

Connection Address: 192.168.254.11

Version B.02.10 [253] SP1 - Advanced | Licenses 32/40

POR
PTB2

17:37:29
15/01/2025

SpectralScan_DEACX26660_15-01-2


[nm]	[mAU]	[nm]	[mAU]	[nm]	[mAU]
200	1428	245	175	290	161
201	1473	**246**	**174**	291	135
202	1517	247	175	292	110
203	1540	248	176	293	88
204	1555	249	182	294	66
205	**1555**	250	189	295	45
206	1547	251	197	296	35
207	1518	252	205	297	23
208	1477	253	215	298	13
209	1416	254	225	299	5
210	1351	255	235	300	-1
211	1285	256	250		
212	1206	257	268		
213	1119	258	286		
214	1040	259	305		
215	965	260	317		
216	895	261	340		
217	826	262	359		
218	772	263	377		
219	714	264	395		
220	648	265	411		
221	597	266	427		
222	548	267	439		
223	521	268	453		
224	478	269	467		
225	450	270	479		
226	426	271	488		
227	409	272	494		
228	393	273	499		
229	377	**274**	**500**		
230	362	275	497		
231	349	276	492		
232	334	277	486		
233	314	278	476		
234	297	279	464		
235	282	280	448		
236	267	281	426		
237	246	282	400		
238	231	283	371		
239	218	284	349		
240	207	285	316		
241	196	286	285		
242	189	287	250		
243	182	288	213		
244	178	289	186		

SAMPLE INFORMATION

Sample Name:	BRANCO 01	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,2	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:19:25 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:38:48 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.108				

SAMPLE INFORMATION

Sample Name:	PCIN 01	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,3	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:22:50 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:39:00 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.017	865747	128335		

SAMPLE INFORMATION

Sample Name:	PCIN 02	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,3	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:26:15 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:39:04 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.018	870087	127718		

SAMPLE INFORMATION

Sample Name:	PCIN 03	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,3	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:29:39 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:39:06 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.017	869592	126889		

SAMPLE INFORMATION

Sample Name:	BRANCO 02	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,2	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:33:04 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:39:07 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.108				

SAMPLE INFORMATION

Sample Name:	PCIN 04	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,3	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:36:29 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:39:09 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.016	868677	125783		

SAMPLE INFORMATION

Sample Name:	PCIN 05	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,3	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:39:53 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:39:10 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.013	867990	124852		

SAMPLE INFORMATION

Sample Name:	PCIN 06	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,3	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:43:17 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:39:12 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.013	867452	124153		

SAMPLE INFORMATION

Sample Name:	BRANCO 03	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,2	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:46:41 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:39:15 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.108				

SAMPLE INFORMATION

Sample Name:	LDUV 01	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,4	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:50:05 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:51:12 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.108				

SAMPLE INFORMATION

Sample Name:	LDUV 02	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,5	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:53:28 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:52:02 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.013	276830	38937		

SAMPLE INFORMATION

Sample Name:	LDUV 03	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,6	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 6:56:51 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:52:03 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.013	576989	81333		

SAMPLE INFORMATION

Sample Name:	LDUV 04	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,7	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 7:00:15 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:52:04 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.010	842812	118112		

SAMPLE INFORMATION

Sample Name:	LDUV 05	Acquired By:	STENICOM
Sample Type:	Unknown	Sample Set Name:	QUALIFICAÇÃO 2025 HPLC19
Vial:	1:A,8	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 7:03:39 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:52:05 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.014	1130088	158994		

SAMPLE INFORMATION

Sample Name:	LIIN 01	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,9	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	0.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 7:06:56 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 1:02:11 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	0.985	1068	92		

SAMPLE INFORMATION

Sample Name:	LIIN 02	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,9	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	5.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 7:10:10 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:52:09 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.001	431340	59245		

SAMPLE INFORMATION

Sample Name:	LIIN 03	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,9	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	10.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 7:13:30 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:52:10 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.012	868189	120913		

SAMPLE INFORMATION

Sample Name:	LIIN 04	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,9	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	15.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 7:16:57 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:52:11 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.018	1300440	182059		

SAMPLE INFORMATION

Sample Name:	LIIN 05	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,9	Acq. Method Set:	HPLC 19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	20.00 ul	Channel Name:	VWD.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Wednesday, January 15, 2025 7:20:30 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 12:52:13 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.026	1734281	242816		

SAMPLE INFORMATION


Sample Name:	RELB	Acquired By:	STENICOM
Sample Type:	Unknown	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,10	Acq. Method Set:	RELB HPLC19 2025
Injection #:	1	Processing Method	Cafeína
Injection Volume:	20.00 ul	Channel Name:	VWD.0.0
Run Time:	20.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 250 nm

Date Acquired: Thursday, January 16, 2025 1:25:22 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 1:51:18 PM Brazil/East -03:00


ALTURA DO PICO



LINHA DE BASE



RUÍDO



Peak Results


	Name	RT	Area	Height	Amount	Units
1	CAFEÍNA	1.036	785374	117392		

SAMPLE INFORMATION

Sample Name:	GRABO_BC	Acquired By:	STENICOM
Sample Type:	Unknown	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,10	Acq. Method Set:	GRABO BC
Injection #:	1	Processing Method	Grad BC
Injection Volume:	0.00 ul	Channel Name:	VWD.0.0
Run Time:	20.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Thursday, January 16, 2025 4:54:30 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 5:34:00 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1		2.266	1593557	12068		
2		5.047	3768351	24615		
3		8.788	6839539	36877		
4		10.147	8763945	49091		
5		13.582	11653188	60292		

SAMPLE INFORMATION

Sample Name:	GRABO_BD	Acquired By:	STENICOM
Sample Type:	Unknown	Sample Set Name:	QUAFICAÇÃO 2025 HPLC19
Vial:	1:A,10	Acq. Method Set:	GRABO BD
Injection #:	1	Processing Method	Grad BD
Injection Volume:	0.00 ul	Channel Name:	VWD.0.0
Run Time:	20.0 Minutes	Proc. Chnl. Descr.:	VWD: Signal A, 273 nm

Date Acquired: Thursday, January 16, 2025 5:15:41 PM America/Buenos_Aires -03:00
Date Processed: Thursday, January 16, 2025 5:36:32 PM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1		1.869	1727168	12134		
2		3.792	3288112	24320		
3		7.542	6792752	36454		
4		10.209	7822846	48462		
5		15.164	12758914	60078		

SAMPLE INFORMATION

Sample Name:	LDRI 1	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	LDRI
Vial:	1:A,1	Acq. Method Set:	LDRI
Injection #:	1	Processing Method	LDRI
Injection Volume:	10.00 ul	Channel Name:	RID.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	RID: RI Signal

Date Acquired: Friday, January 17, 2025 9:53:55 AM America/Buenos_Aires -03:00
Date Processed: Friday, January 17, 2025 10:17:03 AM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	LDRI	0.156	81106	28485		

SAMPLE INFORMATION

Sample Name:	LDRI 2	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	LDRI
Vial:	1:A,2	Acq. Method Set:	LDRI
Injection #:	1	Processing Method	LDRI
Injection Volume:	10.00 ul	Channel Name:	RID.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	RID: RI Signal

Date Acquired: Friday, January 17, 2025 9:57:22 AM America/Buenos_Aires -03:00
Date Processed: Friday, January 17, 2025 10:17:28 AM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	LDRI	0.155	254593	87785		

SAMPLE INFORMATION

Sample Name:	LDRI 3	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	LDRI
Vial:	1:A,3	Acq. Method Set:	LDRI
Injection #:	1	Processing Method	LDRI
Injection Volume:	10.00 ul	Channel Name:	RID.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	RID: RI Signal

Date Acquired: Friday, January 17, 2025 10:00:53 AM America/Buenos_Aires -03:00
Date Processed: Friday, January 17, 2025 10:17:40 AM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	LDRI	0.155	425719	146664		

SAMPLE INFORMATION

Sample Name:	LDRI 4	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	LDRI
Vial:	1:A,4	Acq. Method Set:	LDRI
Injection #:	1	Processing Method	LDRI
Injection Volume:	10.00 ul	Channel Name:	RID.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	RID: RI Signal

Date Acquired: Friday, January 17, 2025 10:04:20 AM America/Buenos_Aires -03:00
Date Processed: Friday, January 17, 2025 10:17:50 AM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results


	Name	RT	Area	Height	Amount	Units
1	LDRI	0.155	596242	205007		

SAMPLE INFORMATION

Sample Name:	LDRI 5	Acquired By:	STENICOM
Sample Type:	Unknow n	Sample Set Name:	LDRI
Vial:	1:A,5	Acq. Method Set:	LDRI
Injection #:	1	Processing Method	LDRI
Injection Volume:	10.00 ul	Channel Name:	RID.0.0
Run Time:	2.0 Minutes	Proc. Chnl. Descr.:	RID: RI Signal

Date Acquired: Friday, January 17, 2025 10:07:46 AM America/Buenos_Aires -03:00
Date Processed: Friday, January 17, 2025 10:17:58 AM Brazil/East -03:00

Auto-Scaled Chromatogram



Peak Results

	Name	RT	Area	Height	Amount	Units
1	LDRI	0.155	853662	294061		