
Chromatogram



Channel & Peak Information Table

Chromatogram Name 200-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,193	597096	67862	100,000	100,000	N/A	479	N/A	1,229	

Chromatogram Name 201-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,193	618631	70431	100,000	100,000	N/A	483	N/A	1,220	

Chromatogram Name 202-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	640305	73427	100,000	100,000	N/A	488	N/A	1,265	

Chromatogram Name 203-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,191	658449	75543	100,000	100,000	N/A	488	N/A	1,251	

Chromatogram Name 204-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,190	677108	77916	100,000	100,000	N/A	490	N/A	1,261	

Chromatogram Name 205-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	685193	78974	100,000	100,000	N/A	491	N/A	1,268	

Chromatogram Name 206-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	681465	79325	100,000	100,000	N/A	503	N/A	1,277	

Chromatogram Name 207-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	675834	78711	100,000	100,000	N/A	505	N/A	1,267	

Chromatogram Name 208-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	664372	77263	100,000	100,000	N/A	503	N/A	1,269	

Chromatogram Name 209-CH1


Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	648559	75645	100,000	100,000	N/A	506	N/A	1,273	

Chromatogram Name 210-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	627192	73203	100,000	100,000	N/A	506	N/A	1,275	

Chromatogram



Channel & Peak Information Table

Chromatogram Name 240-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,190	83762	10145	100,000	100,000	N/A	518	N/A	1,313	

Chromatogram Name 241-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,191	79750	9706	100,000	100,000	N/A	520	N/A	1,313	

Chromatogram Name 242-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	76773	9404	100,000	100,000	N/A	528	N/A	1,332	

Chromatogram Name 243-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,188	72976	8952	100,000	100,000	N/A	528	N/A	1,316	

Chromatogram Name 244-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	70982	9261	100,000	100,000	N/A	605	N/A	1,303	

Chromatogram Name 245-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,185	68012	8936	100,000	100,000	N/A	608	N/A	1,351	

Chromatogram Name 246-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,185	67741	8906	100,000	100,000	N/A	608	N/A	1,360	

Chromatogram Name 247-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,181	69306	9092	100,000	100,000	N/A	607	N/A	1,369	

Chromatogram Name 248-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,184	71381	9329	100,000	100,000	N/A	606	N/A	1,348	

Chromatogram Name 249-CH1


Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,183	74378	9740	100,000	100,000	N/A	607	N/A	1,358	

Chromatogram Name 250-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,185	77664	10155	100,000	100,000	N/A	611	N/A	1,338	

Chromatogram



Channel & Peak Information Table

Chromatogram Name 268-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,186	229353	29133	100,000	100,000	N/A	600	N/A	1,285	

Chromatogram Name 269-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,186	228014	28468	100,000	100,000	N/A	583	N/A	1,227	

Chromatogram Name 270-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,187	232820	29351	100,000	100,000	N/A	597	N/A	1,246	

Chromatogram Name 271-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	236237	29795	100,000	100,000	N/A	596	N/A	1,265	

Chromatogram Name 272-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,187	240242	30352	100,000	100,000	N/A	599	N/A	1,273	

Chromatogram Name 273-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	240354	30397	100,000	100,000	N/A	599	N/A	1,279	

Chromatogram Name 274-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,188	240668	30544	100,000	100,000	N/A	600	N/A	1,291	

Chromatogram Name 275-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,187	237411	30204	100,000	100,000	N/A	603	N/A	1,295	

Chromatogram Name 276-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,185	233740	29816	100,000	100,000	N/A	604	N/A	1,308	

Chromatogram Name 277-CH1


Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,185	228653	29255	100,000	100,000	N/A	608	N/A	1,303	

Chromatogram Name 278-CH1
Sample Name
Channel Name

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,185	221713	28410	100,000	100,000	N/A	608	N/A	1,317	

Chromatogram




Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:19:37
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 16:40:12
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 01-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:19:37
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 16:45:15
Volume	10,0 [µL]
Sample #	3
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 01-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,183	725806	91896	100,000	100,000	N/A	599	N/A	1,299	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:19:37
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 16:50:18
Volume	10,0 [µL]
Sample #	3
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 02-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,183	744630	94213	100,000	100,000	N/A	598	N/A	1,310	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:19:38
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 16:55:20
Volume	10,0 [µL]
Sample #	3
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 03-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,185	735446	92823	100,000	100,000	N/A	598	N/A	1,298	

Creator
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:02:13
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:00:23
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 02-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:19:38
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:05:27
Volume	10,0 [µL]
Sample #	3
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 04-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,185	736102	92724	100,000	100,000	N/A	593	N/A	1,309	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:19:38
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:10:30
Volume	10,0 [µL]
Sample #	3
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 05-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,186	736202	92644	100,000	100,000	N/A	584	N/A	1,287	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:19:38
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:15:35
Volume	10,0 [µL]
Sample #	3
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 06-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,187	735860	92343	100,000	100,000	N/A	580	N/A	1,281	

Creator
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:22:27
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:20:38
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 03-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

Creator
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:27:31
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:25:41
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	LDUV_1-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:32:35
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:30:44
Volume	10,0 [µL]
Sample #	5
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV_2-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	235185	29420	100,000	100,000	N/A	580	N/A	1,275	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:37:39
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:35:50
Volume	10,0 [µL]
Sample #	6
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV_3-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,188	511734	65599	100,000	100,000	N/A	586	N/A	1,292	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:42:47
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:40:58
Volume	10,0 [µL]
Sample #	7
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV_4-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,187	729020	90894	100,000	100,000	N/A	573	N/A	1,289	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 17:47:53
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 17:46:03
Volume	10,0 [µL]
Sample #	8
Project Name	Qualificação HPL 08
Acquisition Time	1,8 [min]
Acquisition Sequence	QPack
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV_5-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,189	966936	120609	100,000	100,000	N/A	575	N/A	1,292	

Creator
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 18:33:49
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 18:31:48
Volume	0,1 [µL]
Sample #	9
Project Name	Qualificação HPL 08
Acquisition Time	2,0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	LIIN 1-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 18:39:01
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 18:36:59
Volume	5,0 [µL]
Sample #	9
Project Name	Qualificação HPL 08
Acquisition Time	2,0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 2-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,185	342355	42346	100,000	100,000	N/A	554	N/A	1,266	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 18:44:15
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 18:42:13
Volume	10,0 [µL]
Sample #	9
Project Name	Qualificação HPL 08
Acquisition Time	2,0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 3-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,193	736507	90702	100,000	100,000	N/A	560	N/A	1,281	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 18:49:35
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 18:47:32
Volume	15,0 [µL]
Sample #	9
Project Name	Qualificação HPL 08
Acquisition Time	2,0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 4-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,201	1140152	140272	100,000	100,000	N/A	566	N/A	1,293	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	23/02/2022 18:54:57
Description	
HPLC System Name	HPL08
Injection Date	23/02/2022 18:52:54
Volume	20,0 [µL]
Sample #	9
Project Name	Qualificação HPL 08
Acquisition Time	2,0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	CAFEÍNA
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 5-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	Peak

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	1	1,207	1543440	190552	100,000	100,000	N/A	577	N/A	1,302	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	24/02/2022 10:15:21
Description	
HPLC System Name	HPL08
Injection Date	24/02/2022 09:52:32
Volume	0,0 [µL]
Sample #	–
Project Name	Qualificação HPL 08
Acquisition Time	20,0 [min]
Acquisition Sequence	
Control Method	Grad_AB
Peak ID Table	Grad AB
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	Grad AB_001-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	(Manual)

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	20%	1	3,721	14607106	93558	6,066	6,986	N/A	N/A	N/A	N/A	
2	40%	1	5,247	29896808	177896	12,416	13,284	N/A	N/A	N/A	N/A	
3	60%	1	8,666	46520261	270971	19,320	20,235	N/A	N/A	N/A	N/A	
4	80%	1	12,849	62343942	355491	25,892	26,546	N/A	N/A	N/A	N/A	
5	100%	1	16,179	87419393	441220	36,306	32,948	N/A	N/A	N/A	N/A	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	24/02/2022 11:04:00
Description	
HPLC System Name	HPL08
Injection Date	24/02/2022 10:38:56
Volume	0,0 [µL]
Sample #	–
Project Name	Qualificação HPL 08
Acquisition Time	20,0 [min]
Acquisition Sequence	
Control Method	Grad_AC
Peak ID Table	Grad AC
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Grad AC_002-CH1
Sample Name	
Channel Name	
Sampling Interval	40 [msec]
Peak Method	(Manual)

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	20%	1	3,363	13779597	88998	5,792	6,728	N/A	N/A	N/A	N/A	
2	40%	1	6,369	28129882	173469	11,823	13,114	N/A	N/A	N/A	N/A	
3	60%	1	9,677	46751860	267303	19,650	20,208	N/A	N/A	N/A	N/A	
4	80%	1	13,173	61587037	351921	25,885	26,605	N/A	N/A	N/A	N/A	
5	100%	1	15,957	87678450	441056	36,851	33,344	N/A	N/A	N/A	N/A	

Creator QPackSA
Reviewer
Approver