
Contour Spectral View




Chromatogram and Results

Injection Details

Injection Name:	**Branco PCIN 01**	*Run Time (min):*	**3,50**
Vial Number:	**BA2**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:00**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
Total:			**0,000**	**0,000**	**0,00**	**0,00**	

Chromatogram and Results

Injection Details

Injection Name:	**PCIN 01**	*Run Time (min):*	**3,50**
Vial Number:	**BA3**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:05**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,683	12,916	157,602	100,00	100,00	n.a.
Total:			**12,916**	**157,602**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**PCIN 02**	*Run Time (min):*	**3,50**
Vial Number:	**BA3**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:10**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,683	12,921	160,627	100,00	100,00	n.a.
Total:			**12,921**	**160,627**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**PCIN 03**	*Run Time (min):*	**3,50**
Vial Number:	**BA3**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:14**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,683	12,920	162,537	100,00	100,00	n.a.
Total:			**12,920**	**162,537**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**Branco PCIN 02**	*Run Time (min):*	**3,50**
Vial Number:	**BA2**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:19**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
Total:			**0,000**	**0,000**	**0,00**	**0,00**	

Chromatogram and Results

Injection Details

Injection Name:	**PCIN 04**	*Run Time (min):*	**3,50**
Vial Number:	**BA3**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:24**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,677	12,919	165,099	100,00	100,00	n.a.
Total:			**12,919**	**165,099**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**PCIN 05**	*Run Time (min):*	**3,50**
Vial Number:	**BA3**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:28**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,677	12,893	165,736	100,00	100,00	n.a.
Total:			**12,893**	**165,736**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**PCIN 06**	*Run Time (min):*	**3,50**
Vial Number:	**BA3**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:33**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,677	12,908	166,610	100,00	100,00	n.a.
Total:			**12,908**	**166,610**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**Branco PCIN 03**	*Run Time (min):*	**3,50**
Vial Number:	**BA2**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:38**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
Total:			**0,000**	**0,000**	**0,00**	**0,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LDUV 01**	*Run Time (min):*	**3,50**
Vial Number:	**BA4**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:43**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,677	0,073	0,910	100,00	100,00	n.a.
Total:			**0,073**	**0,910**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LDUV 02**	*Run Time (min):*	**3,50**
Vial Number:	**BA5**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:47**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,680	4,456	57,671	100,00	100,00	n.a.
Total:			**4,456**	**57,671**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LDUV 03**	*Run Time (min):*	**3,50**
Vial Number:	**BA6**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:52**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,677	8,931	116,388	100,00	100,00	n.a.
Total:			**8,931**	**116,388**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LDUV 04**	*Run Time (min):*	**3,50**
Vial Number:	**BA7**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 18:57**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,677	12,769	166,592	100,00	100,00	n.a.
Total:			**12,769**	**166,592**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LDUV 05**	*Run Time (min):*	**3,50**
Vial Number:	**BA8**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 19:01**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,677	17,294	225,773	100,00	100,00	n.a.
Total:			**17,294**	**225,773**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LIIN 01**	*Run Time (min):*	**3,50**
Vial Number:	**BB1**	*Injection Volume:*	**0,10**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**New Processing Method**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 19:29**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,657	0,114	1,496	100,00	100,00	n.a.
Total:			**0,114**	**1,496**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LIIN 02**	*Run Time (min):*	**3,50**
Vial Number:	**BB1**	*Injection Volume:*	**5,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**New Processing Method**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 19:34**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,667	6,566	85,862	100,00	100,00	n.a.
Total:			**6,566**	**85,862**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LIIN 03**	*Run Time (min):*	**3,50**
Vial Number:	**BB1**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**New Processing Method**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 19:39**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,677	13,173	171,786	100,00	100,00	n.a.
Total:			**13,173**	**171,786**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LIIN 04**	*Run Time (min):*	**3,50**
Vial Number:	**BB1**	*Injection Volume:*	**15,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**New Processing Method**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 19:43**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,680	19,792	258,282	100,00	100,00	n.a.
Total:			**19,792**	**258,282**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**LIIN 05**	*Run Time (min):*	**3,50**
Vial Number:	**BB1**	*Injection Volume:*	**20,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**New Processing Method**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**26/jan/23 19:48**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,690	26,403	340,604	100,00	100,00	n.a.
Total:			**26,403**	**340,604**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**SN 01**	*Run Time (min):*	**15,00**
Vial Number:	**BB2**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**27/jan/23 10:02**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,673	4,396	55,679	100,00	100,00	n.a.
Total:			**4,396**	**55,679**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**SN 01**	*Run Time (min):*	**15,00**
Vial Number:	**BB2**	*Injection Volume:*	**10,00**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**Qualificação QPack HP04**	*Bandwidth:*	**1**
Processing Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**27/jan/23 10:02**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,673	4,396	55,679	100,00	100,00	n.a.
Total:			**4,396**	**55,679**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**GRAD AB**	*Run Time (min):*	**18,00**
Vial Number:	**BA2**	*Injection Volume:*	**0,10**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**GRAD AB**	*Bandwidth:*	**1**
Processing Method:	**New Processing Method(1)**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**27/jan/23 11:07**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results


No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		3,107	130,854	53,408	5,47	6,70	n.a.
2		5,920	284,450	106,662	11,90	13,38	n.a.
3		8,853	437,078	159,778	18,28	20,04	n.a.
4		11,493	624,207	212,679	26,11	26,67	n.a.
5		15,143	914,040	264,874	38,23	33,22	n.a.
Total:			**2390,629**	**797,401**	**100,00**	**100,00**	

Chromatogram and Results

Injection Details

Injection Name:	**GRAD CD**	*Run Time (min):*	**18,00**
Vial Number:	**BA2**	*Injection Volume:*	**0,10**
Injection Type:	**Unknown**	*Channel:*	**UV_VIS_1**
Calibration Level:		*Wavelength:*	**273,0**
Instrument Method:	**GRAD CD**	*Bandwidth:*	**1**
Processing Method:	**New Processing Method(1)**	*Dilution Factor:*	**1,0000**
Injection Date/Time:	**27/jan/23 11:42**	*Sample Weight:*	**1,0000**

Chromatogram



Integration Results

No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Relative Height %	Amount n.a.
1		2,913	124,162	52,188	5,24	6,59	n.a.
2		4,590	293,941	104,165	12,40	13,16	n.a.
3		8,790	414,026	158,665	17,46	20,04	n.a.
4		11,967	609,395	212,175	25,70	26,80	n.a.
5		15,193	929,807	264,577	39,21	33,42	n.a.
Total:			**2371,331**	**791,771**	**100,00**	**100,00**	