
Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-14-11-03-00-01.dx		
Sample name:	BR01		
Description:			
Sample amount:	0.000	**Sample type:**	Blank
Instrument:	CMGA03	**Location:**	1
Injection date:	2025-01-27 16:15:39-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BB	0.1723	141647.8796	475760.7995	100.0000	
		Sum	**141647.8796**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-18-42-03-00-02.dx		
Sample name:	PCIN01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:20:01-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BB	0.1413	141564.9887	481887.9939	98.5416	
0.733	BV	0.0439	31.0800	73.9693	0.0216	
0.873	BB	0.0643	191.3417	368.9281	0.1332	
1.015	BB	0.0437	6.4401	8.5488	0.0045	
1.086	VB	0.2081	1866.2795	2661.3304	1.2991	
		Sum	**143660.1299**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-23-02-03-00-03.dx		
Sample name:	PCIN02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:24:20-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1876	141564.9292	480450.8429	98.5438	
0.733	BV	0.0435	31.0439	73.3669	0.0216	
0.873	BB	0.0593	191.0113	367.0496	0.1330	
1.015	BB	0.0423	6.4327	8.5735	0.0045	
1.086	VV	0.1728	1863.3990	2658.5946	1.2971	
		Sum	**143656.8161**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-27-21-03-00-04.dx		
Sample name:	PCIN03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:28:38-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1838	143176.1853	496710.5660	98.5522	
0.733	BV	0.0352	31.1373	73.8287	0.0214	
0.873	BB	0.0610	192.0855	369.0050	0.1322	
1.015	BB	0.0444	6.4923	8.6423	0.0045	
1.086	VB	0.2081	1873.5978	2684.8914	1.2897	
		Sum	**145279.4981**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-31-40-03-00-05.dx		
Sample name:	BR02		
Description:			
Sample amount:	0.000	**Sample type:**	Blank
Instrument:	CMGA03	**Location:**	1
Injection date:	2025-01-27 16:32:59-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	VV	0.2263	147383.9600	501585.2016	100.0000	
		Sum	**147383.9600**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-36-02-03-00-06.dx		
Sample name:	PCIN04		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:37:19-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1808	144190.7633	496129.7693	98.5431	
0.733	BV	0.0434	31.6568	75.2044	0.0216	
0.873	BB	0.0587	194.7985	375.4766	0.1331	
1.015	BB	0.0410	6.5545	8.7290	0.0045	
1.085	VV	0.1047	1898.7832	2719.2853	1.2977	
		Sum	**146322.5563**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-40-22-03-00-07.dx		
Sample name:	PCIN05		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:41:39-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1505	141775.9290	485187.5151	98.5460	
0.733	BV	0.0418	31.1166	73.9322	0.0216	
0.873	BB	0.0607	191.4200	369.7342	0.1331	
1.015	BB	0.0430	6.7691	8.5762	0.0047	
1.085	VB	0.1742	1862.5448	2673.7869	1.2946	
		Sum	**143867.7795**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-44-42-03-00-08.dx		
Sample name:	PCIN06		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:45:57-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1524	145160.8915	500022.2810	98.5524	
0.733	VV	0.0452	31.6973	75.1438	0.0215	
0.873	BB	0.0563	194.9722	375.3350	0.1324	
1.015	BB	0.0423	6.5976	8.7191	0.0045	
1.085	VV	0.1540	1898.9782	2718.1100	1.2893	
		Sum	**147293.1368**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-48-59-03-00-09.dx		
Sample name:	BR03		
Description:			
Sample amount:	0.000	**Sample type:**	Blank
Instrument:	CMGA03	**Location:**	1
Injection date:	2025-01-27 16:50:15-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.2272	144085.4395	469298.4834	100.0000	
		Sum	**144085.4395**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-53-18-03-00-10.dx		
Sample name:	LDFI01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:54:35-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1538	140938.9766	486965.5052	98.5166	
0.733	BV	0.0453	31.5641	75.1939	0.0221	
0.873	BB	0.0577	194.1061	374.0172	0.1357	
1.015	BB	0.0440	6.5515	8.7079	0.0046	
1.085	VV	0.1541	1889.8863	2714.8526	1.3210	
		Sum	**143061.0846**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-57-37-03-00-11.dx		
Sample name:	LDFI02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:58:54-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1507	140172.5068	478619.1999	98.5229	
0.733	BV	0.0436	31.3036	74.7065	0.0220	
0.872	BB	0.0615	192.3870	371.1172	0.1352	
1.014	BB	0.0419	6.4768	8.5826	0.0046	
1.085	VB	0.2075	1871.3941	2666.3992	1.3153	
		Sum	**142274.0683**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 17-01-57-03-00-12.dx		
Sample name:	LDFI03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 17:03:15-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2024_DefaultMethod.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 09:48:12-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1455	142577.3487	500119.4445	98.5442	
0.733	BV	0.0427	31.3431	75.0429	0.0217	
0.872	BB	0.0586	192.7165	372.4325	0.1332	
1.014	BB	0.0393	6.4683	8.6666	0.0045	
1.085	VB	0.1947	1875.7192	2663.3238	1.2964	
		Sum	**144683.5958**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-31-40-03-00-05.dx		
Sample name:	BR02		
Description:			
Sample amount:	0.000	**Sample type:**	Blank
Instrument:	CMGA03	**Location:**	1
Injection date:	2025-01-27 16:32:59-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	VV	0.2263	147383.9600	501585.2016	100.0000	
		Sum	**147383.9600**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-48-59-03-00-09.dx		
Sample name:	BR03		
Description:			
Sample amount:	0.000	**Sample type:**	Blank
Instrument:	CMGA03	**Location:**	1
Injection date:	2025-01-27 16:50:15-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.2272	144085.4395	469298.4834	100.0000	
		Sum	**144085.4395**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-53-18-03-00-10.dx		
Sample name:	LDFI01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:54:35-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1538	140938.9766	486965.5052	98.5166	
0.733	BV	0.0453	31.5641	75.1939	0.0221	
0.873	BB	0.0577	194.1061	374.0172	0.1357	
1.015	BB	0.0440	6.5515	8.7079	0.0046	
1.085	VV	0.1541	1889.8863	2714.8526	1.3210	
		Sum	**143061.0846**			

Single Injection Report

Agilent

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-57-37-03-00-11.dx		
Sample name:	LDFI02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:58:54-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1507	140172.5068	478619.1999	98.5229	
0.733	BV	0.0436	31.3036	74.7065	0.0220	
0.872	BB	0.0615	192.3870	371.1172	0.1352	
1.014	BB	0.0419	6.4768	8.5826	0.0046	
1.085	VB	0.2075	1871.3941	2666.3992	1.3153	
		Sum	**142274.0683**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 17-01-57-03-00-12.dx		
Sample name:	LDFI03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 17:03:15-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1455	142577.3487	500119.4445	98.5442	
0.733	BV	0.0427	31.3431	75.0429	0.0217	
0.872	BB	0.0586	192.7165	372.4325	0.1332	
1.014	BB	0.0393	6.4683	8.6666	0.0045	
1.085	VB	0.1947	1875.7192	2663.3238	1.2964	
		Sum	**144683.5958**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-44-42-03-00-08.dx		
Sample name:	PCIN06		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:45:57-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1524	145160.8915	500022.2810	98.5524	
0.733	VV	0.0452	31.6973	75.1438	0.0215	
0.873	BB	0.0563	194.9722	375.3350	0.1324	
1.015	BB	0.0423	6.5976	8.7191	0.0045	
1.085	VV	0.1540	1898.9782	2718.1100	1.2893	
		Sum	**147293.1368**			

Single Injection Report

Agilent

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-18-42-03-00-02.dx		
Sample name:	PCIN01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:20:01-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BB	0.1413	141564.9887	481887.9939	98.5416	
0.733	BV	0.0439	31.0800	73.9693	0.0216	
0.873	BB	0.0643	191.3417	368.9281	0.1332	
1.015	BB	0.0437	6.4401	8.5488	0.0045	
1.086	VB	0.2081	1866.2795	2661.3304	1.2991	
		Sum	**143660.1299**			

Single Injection Report

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-23-02-03-00-03.dx		
Sample name:	PCIN02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:24:20-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1876	141564.9292	480450.8429	98.5438	
0.733	BV	0.0435	31.0439	73.3669	0.0216	
0.873	BB	0.0593	191.0113	367.0496	0.1330	
1.015	BB	0.0423	6.4327	8.5735	0.0045	
1.086	VV	0.1728	1863.3990	2658.5946	1.2971	
		Sum	**143656.8161**			

Single Injection Report

Agilent

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-27-21-03-00-04.dx		
Sample name:	PCIN03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:28:38-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1838	143176.1853	496710.5660	98.5522	
0.733	BV	0.0352	31.1373	73.8287	0.0214	
0.873	BB	0.0610	192.0855	369.0050	0.1322	
1.015	BB	0.0444	6.4923	8.6423	0.0045	
1.086	VB	0.2081	1873.5978	2684.8914	1.2897	
		Sum	**145279.4981**			

Single Injection Report

Agilent

Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-36-02-03-00-06.dx		
Sample name:	PCIN04		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:37:19-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		




Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1808	144190.7633	496129.7693	98.5431	
0.733	BV	0.0434	31.6568	75.2044	0.0216	
0.873	BB	0.0587	194.7985	375.4766	0.1331	
1.015	BB	0.0410	6.5545	8.7290	0.0045	
1.085	VV	0.1047	1898.7832	2719.2853	1.2977	
		Sum	**146322.5563**			


Single Injection Report


Data file:	C:\CDSProjects\Agilent\Results\CMGA003\5 - 2025\1 - Janeiro\CMGA032025-01-27 16-14-00-03-00.rslt\2025-01-27 16-40-22-03-00-07.dx		
Sample name:	PCIN05		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CMGA03	**Location:**	2
Injection date:	2025-01-27 16:41:39-03:00	**Injection:**	1 of 1
Acq. method:	QPACK QUAL PCIN BACK-2025.amx	**Injection volume:**	1.000 µL
Analysis method:	QPACK 2025.pmx	**Acq. operator:**	SYSTEM (SYSTEM)
Last changed:	2025-01-28 10:02:52-03:00		



Signal: FID2B

RT [min]	Type	Width [min]	Area	Height	Area%	Name
0.472	BV	0.1505	141775.9290	485187.5151	98.5460	
0.733	BV	0.0418	31.1166	73.9322	0.0216	
0.873	BB	0.0607	191.4200	369.7342	0.1331	
1.015	BB	0.0430	6.7691	8.5762	0.0047	
1.085	VB	0.1742	1862.5448	2673.7869	1.2946	
		Sum	**143867.7795**			


Single Sample Analysis
Priority Sample
Run
Run Queue
Online Signals
Separate
Agilent 7890B
Front Signal
Back Signal
Back Signal
Free zoom
DESVIO DE LINHA BASE 20 MIN DE 7,30pA A 6,95pA
TOTALIZANDO 0,35pA
pA
Signal Selection
CMGA03 - Idle


Single Sample Analysis
Priority Sample
Run
Run Queue
Online Signals
Separate
Agilent 7890B
Front Signal
Back Signal
Back Signal
Free zoom
RUÍDO DE 7,08pA A 7,02pA
TOTALIZANDO 0,06pA
pA
Signal Selection