
Peak Find - EDUV PDA_1,7467.JWS



[Comments]

User	QPackSA
Division	---
Company	Valer Laboratórios
Creation date	07/05/2021 11:12

[Detailed Information]

Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	200 nm
End	300 nm
Data interval	1 nm
Data points	101


[Measurement Information]

Instrument Name	MD-4010
Model Name	MD-4010
Serial No.	A011361665
ROM version	V1.01.22

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	273	0.034337	2	245	0.009636
3	206	0.091558			

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 14:21:58
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 11:48:11
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco_01-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 14:21:59
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 11:51:24
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN_01-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,725	811531	99553	100,000	100,000	N/A	984	N/A	1,229	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 14:21:59
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 11:54:39
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN_02-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,702	810252	100094	100,000	100,000	N/A	972	N/A	1,229	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 14:22:00
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 11:57:53
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN_03-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,701	808133	99872	100,000	100,000	N/A	971	N/A	1,238	

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 14:22:00
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:01:06
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco_02-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 14:22:01
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:04:20
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN_04-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,696	811898	100384	100,000	100,000	N/A	967	N/A	1,242	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 14:22:01
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:07:35
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN_05-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,692	810512	100098	100,000	100,000	N/A	961	N/A	1,247	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 14:22:01
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:10:49
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN_06-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,695	810488	99798	100,000	100,000	N/A	956	N/A	1,250	

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 14:22:02
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:14:02
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco_03-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:28
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:17:16
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	LDUV 01-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:31
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:20:30
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 02-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,697	270486	33219	100,000	100,000	N/A	947	N/A	1,242	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:31
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:23:45
Volume	10,0 [µL]
Sample #	3
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 03-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,692	521225	62251	100,000	100,000	N/A	879	N/A	1,283	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:32
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:26:58
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 04-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,694	809693	98821	100,000	100,000	N/A	936	N/A	1,259	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:32
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:30:12
Volume	10,0 [µL]
Sample #	5
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 05-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,697	1036402	125689	100,000	100,000	N/A	925	N/A	1,276	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:32
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:32:55
Volume	0,1 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 01-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,695	7186	883	100,000	100,000	N/A	882	N/A	1,293	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:33
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:35:53
Volume	5,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 02-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,707	399178	47544	100,000	100,000	N/A	906	N/A	1,302	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:33
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:39:06
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 03-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,716	809772	96994	100,000	100,000	N/A	926	N/A	1,283	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:33
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:42:35
Volume	15,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 04-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,725	1225709	146274	100,000	100,000	N/A	931	N/A	1,267	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	07/05/2021 13:50:34
Description	
HPLC System Name	HPL19
Injection Date	07/05/2021 12:46:18
Volume	20,0 [µL]
Sample #	4
Project Name	Qualificação Operacional
Acquisition Time	2,3 [min]
Acquisition Sequence	Qualificação
Control Method	Qualificação
Peak ID Table	Cafeína
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 05-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Peak Method Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1,731	1645583	194927	100,000	100,000	N/A	924	N/A	1,264	

Creator QPackSA
Reviewer
Approver

Chromatogram




Channel & Peak Information Table

Chromatogram Name GRAD-AB_001-CH5
Sample Name
Channel Name 273,0nm

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	20%	5	4,420	12209274	85458	5,003	6,313	N/A	N/A	N/A	N/A	
2	40%	5	7,463	29821039	177874	12,219	13,140	N/A	N/A	N/A	N/A	
3	60%	5	10,443	44530645	271102	18,247	20,027	N/A	N/A	N/A	N/A	
4	80%	5	13,397	61710580	364758	25,286	26,945	N/A	N/A	N/A	N/A	
5	100%	5	16,527	95775160	454503	39,245	33,575	N/A	N/A	N/A	N/A	

Chromatogram



Channel & Peak Information Table

Chromatogram Name GRAD-CD_001-CH5
Sample Name
Channel Name 273,0nm

#	Peak Name	CH	tR [min]	Area [μV·sec]	Height [μV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	20%	5	4,473	12700203	86916	5,190	6,400	N/A	N/A	N/A	N/A	
2	40%	5	7,373	28541850	178942	11,664	13,175	N/A	N/A	N/A	N/A	
3	60%	5	10,227	44439415	272161	18,161	20,039	N/A	N/A	N/A	N/A	
4	80%	5	13,517	65230556	365475	26,657	26,909	N/A	N/A	N/A	N/A	
5	100%	5	16,567	93791593	454673	38,329	33,477	N/A	N/A	N/A	N/A	