


Chromeleon - PC-U3000\2019\Qualificação\EDUV #1 EDUV (UV_VIS_1) - [PC-U3000\2019\Qualificação\EDUV #1 EDUV (UV_VIS_1)]
File Edit View Workspace Qualification Window Help
EDUV #1
EDUV
UV_VIS_1
mAU
WVL:273 nm
8 - 2,743
SB1
SB2
min
Peak #8 100% at 2.74 min
%
50% at 2.69 min: 1000.00
-50% at 2.81 min: 1000.00
205.1
273.6
nm
Zoom from 243,0 nm, 19 % to 245,1 nm, -15 %
rute
PT
11:26
26/04/2019

1 BR PCIN 001


Sample Name:	**BR PCIN 001**	*Injection Volume:*	**10,0**
Vial Number:	**BA2**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 11:35**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
Total:			0,000	0,000	0,00	0,000	

2 PCIN 001


Sample Name:	**PCIN 001**	*Injection Volume:*	**10,0**
Vial Number:	**BA3**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 11:42**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,59	n.a.	93,833	14,322	100,00	n.a.	BMB
Total:			93,833	14,322	100,00	0,000	

3 PCIN 002


Sample Name:	**PCIN 002**	*Injection Volume:*	**10,0**
Vial Number:	**BA3**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 11:49**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,56	n.a.	95,078	14,312	100,00	n.a.	BMB
Total:			95,078	14,312	100,00	0,000	

4 PCIN 003


Sample Name:	**PCIN 003**	*Injection Volume:*	**10,0**
Vial Number:	**BA3**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 11:56**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,53	n.a.	96,165	14,336	100,00	n.a.	BMB
Total:			96,165	14,336	100,00	0,000	

5 BR PCIN 002


Sample Name:	**BR PCIN 002**	*Injection Volume:*	**10,0**
Vial Number:	**BA2**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 12:03**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
Total:			0,000	0,000	0,00	0,000	

6 PCIN 004


Sample Name:	**PCIN 004**	*Injection Volume:*	**10,0**
Vial Number:	**BA3**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 12:09**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,47	n.a.	100,001	14,363	100,00	n.a.	BMB
Total:			100,001	14,363	100,00	0,000	

7 PCIN 005


Sample Name:	**PCIN 005**	*Injection Volume:*	**10,0**
Vial Number:	**BA3**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 12:16**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,43	n.a.	101,620	14,392	100,00	n.a.	BMB
Total:			101,620	14,392	100,00	0,000	

8 PCIN 006


Sample Name:	**PCIN 006**	*Injection Volume:*	**10,0**
Vial Number:	**BA3**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 12:23**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,40	n.a.	102,965	14,390	100,00	n.a.	BMB
Total:			102,965	14,390	100,00	0,000	

9 BR PCIN 003


Sample Name:	**BR PCIN 003**	*Injection Volume:*	**10,0**
Vial Number:	**BA2**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 12:30**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
Total:			0,000	0,000	0,00	0,000	

1 LDUV 001


Sample Name:	**LDUV 001**	*Injection Volume:*	**10,0**
Vial Number:	**BA4**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 12:59**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,53	n.a.	0,094	0,014	100,00	n.a.	BMB
Total:			0,094	0,014	100,00	0,000	

2 LDUV 002


Sample Name:	**LDUV 002**	*Injection Volume:*	**10,0**
Vial Number:	**BA5**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 13:07**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,21	n.a.	39,500	4,870	100,00	n.a.	BMB
Total:			39,500	4,870	100,00	0,000	

3 LDUV 003


Sample Name:	**LDUV 003**	*Injection Volume:*	**10,0**
Vial Number:	**BA6**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 13:15**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,18	n.a.	69,816	9,483	100,00	n.a.	BMB
Total:			69,816	9,483	100,00	0,000	

4 LDUV 004


Sample Name:	**LDUV 004**	*Injection Volume:*	**10,0**
Vial Number:	**BA7**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 13:24**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,14	n.a.	105,653	13,997	100,00	n.a.	BMB*
Total:			105,653	13,997	100,00	0,000	

5 LDUV 005


Sample Name:	**LDUV 005**	*Injection Volume:*	**10,0**
Vial Number:	**BA8**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 13:32**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,10	n.a.	143,719	18,859	100,00	n.a.	MB*
Total:			143,719	18,859	100,00	0,000	

1 LIIN 001


Sample Name:	**LIIN 001**	*Injection Volume:*	**0,1**
Vial Number:	**BB1**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 13:46**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,04	n.a.	1,072	0,131	100,00	n.a.	BMB
Total:			1,072	0,131	100,00	0,000	

2 LIIN 002


Sample Name:	**LIIN 002**	*Injection Volume:*	**5,0**
Vial Number:	**BB1**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 13:54**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	2,01	n.a.	54,019	7,086	100,00	n.a.	BMB
Total:			54,019	7,086	100,00	0,000	

3 LIIN 003


Sample Name:	**LIIN 003**	*Injection Volume:*	**10,0**
Vial Number:	**BB1**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 14:03**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	1,98	n.a.	106,811	14,098	100,00	n.a.	BMB
Total:			106,811	14,098	100,00	0,000	

4 LIIN 004


Sample Name:	**LIIN 004**	*Injection Volume:*	**15,0**
Vial Number:	**BB1**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 14:11**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	1,97	n.a.	159,612	21,175	100,00	n.a.	BMB
Total:			159,612	21,175	100,00	0,000	

5 LIIN 005


Sample Name:	**LIIN 005**	*Injection Volume:*	**20,0**
Vial Number:	**BB1**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**Qualificação**	*Bandwidth:*	**4**
Quantif. Method:	**Qualificação**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 14:19**	*Sample Weight:*	**1,0000**
Run Time (min):	**5,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	1,96	n.a.	224,355	28,265	100,00	n.a.	BMB
Total:			224,355	28,265	100,00	0,000	

1 GRAD AB


Sample Name:	**GRAD AB**	*Injection Volume:*	**0,0**
Vial Number:	**BA2**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**GRAD AB**	*Bandwidth:*	**4**
Quantif. Method:	**GRAD**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 15:07**	*Sample Weight:*	**1,0000**
Run Time (min):	**18,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	1,96	n.a.	68,505	92,725	3,03	n.a.	BM *
2	4,96	n.a.	135,955	301,030	9,85	n.a.	M *
3	8,10	n.a.	203,862	525,491	17,20	n.a.	M *
4	11,18	n.a.	271,294	737,085	24,12	n.a.	M *
5	14,34	n.a.	338,053	974,506	31,89	n.a.	M *
6	14,59	n.a.	0,000	0,000	0,00	n.a.	Ru*
7	15,12	n.a.	338,084	425,025	13,91	n.a.	MB*
Total:			1355,754	3055,862	100,00	0,000	

1 GRAD CD

Sample Name:	**GRAD CD**	*Injection Volume:*	**0,0**
Vial Number:	**BA2**	*Channel:*	**UV_VIS_1**
Sample Type:	**unknown**	*Wavelength:*	**273.0**
Control Program:	**GRAD CD**	*Bandwidth:*	**4**
Quantif. Method:	**GRAD**	*Dilution Factor:*	**1,0000**
Recording Time:	**26/4/2019 15:27**	*Sample Weight:*	**1,0000**
Run Time (min):	**18,00**	*Sample Amount:*	**1,0000**



No.	Ret.Time min	Peak Name	Height mAU	Area mAU*min	Rel.Area %	Amount	Type
1	1,90	n.a.	70,935	166,354	5,50	n.a.	BM *
2	5,01	n.a.	134,763	231,417	7,65	n.a.	M *
3	7,86	n.a.	201,420	498,302	16,47	n.a.	M *
4	11,17	n.a.	268,379	741,345	24,50	n.a.	M *
5	15,19	n.a.	335,632	1388,143	45,88	n.a.	MB*
Total:			1011,129	3025,560	100,00	0,000	