INSTRUMENT PERFORMANCE REPORT

Manufacturers Name : Polymer Laboratories Ltd.

Manufacturers Address : Essex Road
Church Stretton
Shropshire
SY6 6AX

Instrument Name : PL-ELS 1000

Serial Number : 003_001

The following functional characteristics of the PL-ELS 1000 instrument have all been individually tested for conformance with our internal specifications.

Temperature Response : PASSED

Both the nebuliser and evaporation heaters have successfully reached the desired temperature within the expected time period.

Temperature Accuracy : PASSED

Both the nebuliser and evaporator were successfully maintaining a constant temperature at the desired setpoints.

Baseline Stability : PASSED

The output of your PL-ELS 1000 instrument has shown to satisfy the test criteria of the baseline test.

Mass Flow Response : PASSED

The mass flow controller was successfully reaching and maintaining the desired gas flow rates set.

Report Date : quarta-feira, mar‡o 29, 2023

Test Technician : _____