

MicroLab
MicroLab
User: Equiplex\Desenvolvimento\karita souza
Status: Ready
Method: Polietileno_MP
21CFR11 Enabled
Current Instrument Type: Cary 630
Current Sampling Technology: Diamond ATR 1-Bounce
Instrument Status
Energy: 24734
Battery: On AC power
Source: 1.896 / 2.703
Laser: 7535
Gain Adjust
Temperatures (°C)
Detector: 40.022
CPU: 45.75
IR board: 51.213
Power: 46.25
Block: 42.625
Version Information
Version: 5.8.2648.0
DLL Version: 5.8.88.0
Firmware Version: 1.96.00.0
Instrument Serial Number:
MY2451CU06
Resetting the Clean validity will cause the next clean run to be stored as the new 'Good' Clean run.
Reset Clean Validity
For methods that do not require a new background on every sample, the 'Reset Background Validity' button will force a new background to be collected before the next sample collection occurs.
Reset Background Validity
Date Time format configuration (for generation of DateTime in files' names) :
Fixed: yyyy-MM-ddThh-mm-ss
Live Display Spectrum Preview Prior to Sample Collection
Home
< Back
Agilent Technologies

This Report was generated under a 21 CFR 11 Environment

System Information:

Sample ID: 01 numero de onda

System Status: Good

File Location: HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\ 01 numero de onda.a2r

PC/Workstation ID: GHE-8180

Software Version:MicroLab PC(21CFR11): 5.8.2648.0

FW Version:1.96.00.0

Data Information:

Method Name: HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|ATR EXATIDAO NUMERO DE ONDA.a2m

User: Equiplex\Desenvolvimento\karita souza

Date/Time: 03/31/2026 9:31:02 AM

Instrument Model: Cary630

Instrument S/N: MY2451CU06

Reporting Information:

Report Template: Spectrum.docx


Report Generated By: Equiplex\Desenvolvimento\karita souza

Report Generation Date: 03/31/2026 10:14:33 AM
Report Generation GMT Date:03/31/2026 1:14:33 PM

Sample ID:01 numero de onda

Method Name:HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|ATR EXATIDAO NUMERO DE ONDA.a2m

Sample Scans:16
Background Scans:16
Resolution:2
System Status:Good
User:Equiplex\Desenvolvimento\karita souza
Date/Time:03/31/2026 9:31:02 AM
Range:4000 - 650
Apodization:Happ-Genzel
File Location:HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\01 numero de onda.a2r



Peak Number	Wavenumber (cm^{-1})	Intensity
1	711.0	0.379
2	1796.0	0.842
3	2851.5	0.579
4	2921.9	0.487

E-SIGNATURES

Signature

This Report was generated under a 21 CFR 11 Environment

System Information:
Sample ID: RESOLUCAO

System Status: Good

File Location: HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\ RESOLUCAO.a2r

PC/Workstation ID: GHE-8180

Software Version:MicroLab PC(21CFR11): 5.8.2648.0

FW Version:1.96.00.0

Data Information:
Method Name: HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|RESOLUCAO.a2m

User: Equiplex\Desenvolvimento\karita souza

Date/Time: 03/31/2026 10:12:09 AM

Instrument Model: Cary630

Instrument S/N: MY2451CU06

Reporting Information:
Report Template: Spectrum.docx

Report Generated By: Equiplex\Desenvolvimento\karita souza


Report Generation Date: 03/31/2026 10:18:38 AM
Report Generation GMT Date:03/31/2026 1:18:38 PM

Sample ID:RESOLUCAO

Sample Scans:16
Background Scans:16
Resolution:2

Method Name:HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|RESOLUCAO.a2m
User:Equiplex\Desenvolvimento\karita souza
Date/Time:03/31/2026 10:12:09 AM
Range:4000 - 650

System Status:Good Apodization:Happ-Genzel
File Location:HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\RESOLUCAO.a2r



Peak Number	Wavenumber (cm^{-1})	Intensity
1	850.1	72.097
2	871.3	37.691

E-SIGNATURES

Signature

This Report was generated under a 21 CFR 11 Environment

System Information:

Sample ID: REP NUMERO DE ONDA

System Status: Good

File Location: HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\ REP NUMERO DE ONDA.a2r

PC/Workstation ID: GHE-8180

Software Version:MicroLab PC(21CFR11): 5.8.2648.0

FW Version:1.96.00.0

Data Information:

Method Name: HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|ATR REP NUMERO DE ONDAa2m.a2m

User: Equiplex\Desenvolvimento\karita souza

Date/Time: 03/31/2026 10:23:33 AM

Instrument Model: Cary630

Instrument S/N: MY2451CU06

Reporting Information:

Report Template: Spectrum.docx

Report Generated By: Equiplex\Desenvolvimento\karita souza

Report Generation Date: 03/31/2026 10:56:45 AM

Report Generation GMT Date:03/31/2026 1:56:45 PM

Sample ID:REP NUMERO DE ONDA

Sample Scans:16


Background Scans:16

Method Name:HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|ATR REP NUMERO DE ONDAa2m.a2m

User:Equiplex\Desenvolvimento\karita souza

Date/Time:03/31/2026 10:23:33 AM

Resolution:2 Range:4000 - 650
System Status:Good Apodization:Happ-Genzel
File Location:HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\REP NUMERO DE ONDA.a2r



Peak Number	Wavenumber (cm^{-1})	Intensity
1	712.6	49.459
2	1796.0	86.796
3	2853.2	64.533
4	2921.9	55.382

E-SIGNATURES

Signature

This Report was generated under a 21 CFR 11 Environment

System Information:

Sample ID: REP NUMERO DE ONDA 2a2r

System Status: Good

File Location: HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\ REP NUMERO DE ONDA 2a2r.a2r

PC/Workstation ID: GHE-8180

Software Version:MicroLab PC(21CFR11): 5.8.2648.0

FW Version:1.96.00.0

Data Information:

Method Name: HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|ATR REP NUMERO DE ONDAa2m.a2m

User: Equiplex\Desenvolvimento\karita souza

Date/Time: 03/31/2026 10:27:56 AM

Instrument Model: Cary630

Instrument S/N: MY2451CU06

Reporting Information:

Report Template: Spectrum.docx

Report Generated By: Equiplex\Desenvolvimento\karita souza


Report Generation Date: 03/31/2026 10:31:39 AM
Report Generation GMT Date:03/31/2026 1:31:39 PM

Sample ID:REP NUMERO DE ONDA 2a2r

Sample Scans:16
Background Scans:16

Method Name:HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|ATR REP NUMERO DE ONDAa2m.a2m
User:Equiplex\Desenvolvimento\karita souza
Date/Time:03/31/2026 10:27:56 AM

Resolution:2 Range:4000 - 650
System Status:Good Apodization:Happ-Genzel
File Location:HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\REP NUMERO DE ONDA 2a2r.a2r



Peak Number	Wavenumber (cm^{-1})	Intensity
1	712.6	49.149
2	1796.0	88.936
3	2853.2	65.216
4	2921.9	55.508

E-SIGNATURES

Signature

This Report was generated under a 21 CFR 11 Environment

System Information:

Sample ID: atr rep foto

System Status: Good

File Location: HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\ atr rep foto.a2r

PC/Workstation ID: GHE-8180

Software Version:MicroLab PC(21CFR11): 5.8.2648.0

FW Version:1.96.00.0

Data Information:

Method Name: HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|REP FOTOMETRCA.a2m

User: Equiplex\Desenvolvimento\KARITA SOUZA

Date/Time: 03/31/2026 3:56:24 PM

Instrument Model: Cary630

Instrument S/N: MY2451CU06

Reporting Information:

Report Template: Spectrum.docx

Report Generated By: Equiplex\Desenvolvimento\KARITA SOUZA

Report Generation Date: 03/31/2026 3:57:52 PM

Report Generation GMT Date:03/31/2026 6:57:52 PM


Sample ID:atr rep foto

Sample Scans:16

Method Name:HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|REP FOTOMETRCA.a2m

User:Equiplex\Desenvolvimento\KARITA SOUZA

Background Scans:16 Date/Time:03/31/2026 3:56:24 PM
Resolution:2 Range:4000 - 650
System Status:Good Apodization:Happ-Genzel
File Location:HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\atr rep foto.a2r



Peak Number	Wavenumber (cm^{-1})	Intensity
1	712.6	57.616
2	1796.0	87.643
3	2853.2	70.524
4	2921.9	64.397

E-SIGNATURES

Signature

This Report was generated under a 21 CFR 11 Environment

System Information:

Sample ID: atr rep foto_

System Status: Good

File Location: HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\ atr rep foto_.a2r

PC/Workstation ID: GHE-8180

Software Version:MicroLab PC(21CFR11): 5.8.2648.0

FW Version:1.96.00.0

Data Information:

Method Name: HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|REP FOTOMETRCA.a2m

User: Equiplex\Desenvolvimento\KARITA SOUZA

Date/Time: 03/31/2026 3:56:24 PM

Instrument Model: Cary630

Instrument S/N: MY2451CU06

Reporting Information:

Report Template: Spectrum.docx

Report Generated By: Equiplex\Desenvolvimento\KARITA SOUZA

Report Generation Date: 03/31/2026 3:59:46 PM

Report Generation GMT Date:03/31/2026 6:59:46 PM


Sample ID:atr rep foto_

Sample Scans:16

Method Name:HEG-HPLC\SQLEXPRESS|VAIMDB_SystemCheckMethods(000)|REP FOTOMETRCA.a2m

User:Equiplex\Desenvolvimento\KARITA SOUZA


Background Scans:16 Date/Time:03/31/2026 3:56:24 PM
Resolution:2 Range:4000 - 650
System Status:Good Apodization:Happ-Genzel
File Location:HEG-HPLC\SQLEXPRESS|VAIMDB_Public(000)|Public Results\atr rep foto_.a2r



Peak Number	Wavenumber (cm^{-1})	Intensity
1	712.6	57.616
2	1396.7	30.208
3	1796.0	87.643
4	2853.2	70.524
5	2921.9	64.397

E-SIGNATURES

Signature


MicroLab
MicroLab
User: Equiplex\Desenvolvimento\karita souza
Status: Ready
Method: StabilityTest
21CFR11 Enabled
System Check Results:
Name
Value
Mean
Std Dev
Co-Var
Min
Max
Stability Point 1
100.0030
100.0890
Stability Point 2
100.1763
102.6953
Home
Next >
Agilent Technologies