


Certificado N°: 221/21

- **DETECTOR UV-Vis - WATERS 2487 - C99487796M:**








QPack Soluções Analíticas Ltda.
Rua Ema Gazzi Magnusson, 138 – Comercial Vitoria Martini – Indaiatuba – SP – CEP 13347-630
Site: www.qpack.com.br - Fone: 55 (19) 2660-1110
E-mail: qpack@qpack.com.br

SAMPLE INFORMATION

Sample Name:	Branco 1	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 12:47:30 BRT		
Date Processed:	21/09/2021 14:06:14 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,599				

SAMPLE INFORMATION

Sample Name:	PCIN 1	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 12:51:42 BRT		
Date Processed:	21/09/2021 14:06:21 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,537	375002	54395		

SAMPLE INFORMATION

Sample Name:	PCIN 2	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 12:55:49 BRT		
Date Processed:	21/09/2021 14:06:27 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,565	377920	53857		

SAMPLE INFORMATION

Sample Name:	PCIN 3	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 12:59:56 BRT		
Date Processed:	21/09/2021 14:06:32 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,571	378105	53726		

SAMPLE INFORMATION

Sample Name:	Branco 2	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 13:04:08 BRT		
Date Processed:	21/09/2021 14:06:38 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,599				

SAMPLE INFORMATION

Sample Name:	PCIN 4	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 13:08:18 BRT		
Date Processed:	21/09/2021 14:06:43 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,579	380074	53323		

SAMPLE INFORMATION

Sample Name:	PCIN 5	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 13:12:27 BRT		
Date Processed:	21/09/2021 14:06:48 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,584	384426	54158		

SAMPLE INFORMATION

Sample Name:	PCIN 6	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 13:16:37 BRT		
Date Processed:	21/09/2021 14:06:54 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,584	380523	53220		

SAMPLE INFORMATION

Sample Name:	Branco 3	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 13:20:47 BRT		
Date Processed:	21/09/2021 14:06:57 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,599				

SAMPLE INFORMATION

Sample Name:	LDUV 1	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 14:06:34 BRT		
Date Processed:	21/09/2021 14:30:45 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,599				

SAMPLE INFORMATION

Sample Name:	LDUV 2	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 14:10:46 BRT		
Date Processed:	21/09/2021 14:30:50 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,540	161623	23129		

SAMPLE INFORMATION

Sample Name:	LDUV 3	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	3	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 14:14:57 BRT		
Date Processed:	21/09/2021 14:30:55 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,539	337518	48196		

SAMPLE INFORMATION

Sample Name:	LDUV 4	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	4	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 14:19:08 BRT		
Date Processed:	21/09/2021 14:31:00 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,533	487425	70012		

SAMPLE INFORMATION

Sample Name:	LDUV 5	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	5	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	10,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 14:23:20 BRT		
Date Processed:	21/09/2021 14:31:04 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,546	647014	93794		

SAMPLE INFORMATION

Sample Name:	LIIN 1	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	0,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 15:30:19 BRT		
Date Processed:	21/09/2021 15:51:31 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,599				

SAMPLE INFORMATION

Sample Name:	LIIN 2	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	225,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 15:34:27 BRT		
Date Processed:	21/09/2021 15:51:36 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,563	394982	53632		

SAMPLE INFORMATION

Sample Name:	LIIN 3	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	450,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 15:38:37 BRT		
Date Processed:	21/09/2021 15:51:41 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,581	807990	120089		

SAMPLE INFORMATION

Sample Name:	LIIN 4	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	675,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 15:42:46 BRT		
Date Processed:	21/09/2021 15:51:46 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,589	1229136	186877		

SAMPLE INFORMATION

Sample Name:	LIIN 5	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	900,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	21/09/2021 15:46:54 BRT		
Date Processed:	21/09/2021 15:51:52 BRT		

Auto-Scaled Chromatogram



Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,597	1670140	249273		