

Certificado N°: 222/21

- **DETECTOR UV-Vis - WATERS 2487 – B98487184M:**






QPack Soluções Analíticas Ltda.
Rua Ema Gazzi Magnusson, 138 – Comercial Vitoria Martini – Indaiatuba – SP – CEP 13347-630
Site: www.qpack.com.br - Fone: 55 (19) 2660-1110
E-mail: qpack@qpack.com.br

SAMPLE INFORMATION

Sample Name:	Branco 1	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:13:57 BRT		
Date Processed:	22/09/2021 12:49:09 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,671				

SAMPLE INFORMATION

Sample Name:	PCIN 1	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:18:47 BRT		
Date Processed:	22/09/2021 12:49:17 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,675	4295706	583515		

SAMPLE INFORMATION

Sample Name:	PCIN 2	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:23:37 BRT		
Date Processed:	22/09/2021 12:49:23 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,668	4293880	591966		

SAMPLE INFORMATION

Sample Name:	PCIN 3	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:28:26 BRT		
Date Processed:	22/09/2021 12:49:30 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,663	4291733	598137		

SAMPLE INFORMATION

Sample Name:	Branco 2	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:33:17 BRT		
Date Processed:	22/09/2021 12:49:36 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,671				

SAMPLE INFORMATION

Sample Name:	PCIN 4	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:38:07 BRT		
Date Processed:	22/09/2021 12:49:47 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,657	4296284	601279		

SAMPLE INFORMATION

Sample Name:	PCIN 5	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:42:55 BRT		
Date Processed:	22/09/2021 12:49:53 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,655	4290584	609008		

SAMPLE INFORMATION

Sample Name:	PCIN 6	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:47:46 BRT		
Date Processed:	22/09/2021 12:49:59 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,656	4293102	608073		

SAMPLE INFORMATION

Sample Name:	Branco 3	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	PCIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:52:39 BRT		
Date Processed:	22/09/2021 12:50:05 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,671				

SAMPLE INFORMATION

Sample Name:	LDUV 1	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 11:58:29 BRT		
Date Processed:	22/09/2021 12:50:11 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,671				

SAMPLE INFORMATION

Sample Name:	LDUV 2	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	2	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 12:03:21 BRT		
Date Processed:	22/09/2021 12:50:17 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,678	1847969	283699		

SAMPLE INFORMATION

Sample Name:	LDUV 3	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	3	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 12:08:12 BRT		
Date Processed:	22/09/2021 12:50:22 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,673	3755460	537816		

SAMPLE INFORMATION

Sample Name:	LDUV 4	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	4	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 12:13:03 BRT		
Date Processed:	22/09/2021 12:50:28 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,669	5441345	787382		

SAMPLE INFORMATION

Sample Name:	LDUV 5	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LDUV
Vial:	5	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 12:17:54 BRT		
Date Processed:	22/09/2021 12:50:33 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,669	7116418	1110689		

SAMPLE INFORMATION

Sample Name:	LIIN 1	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	0,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 12:50:05 BRT		
Date Processed:	22/09/2021 13:18:13 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,671				

SAMPLE INFORMATION

Sample Name:	LIIN 2	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	25,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 12:54:24 BRT		
Date Processed:	22/09/2021 13:18:18 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,616	1360040	210469		

SAMPLE INFORMATION

Sample Name:	LIIN 3	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	50,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 12:58:53 BRT		
Date Processed:	22/09/2021 13:18:23 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,639	2744144	427285		

SAMPLE INFORMATION

Sample Name:	LIIN 4	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	75,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 13:03:30 BRT		
Date Processed:	22/09/2021 13:18:29 BRT		

Auto-Scaled Chromatogram




Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,653	4095947	642686		

SAMPLE INFORMATION

Sample Name:	LIIN 5	Acquired By:	System
Sample Type:	Unknown	Sample Set Name:	LIIN
Vial:	1	Acq. Method Set:	Qualificação
Injection #:	1	Processing Method:	Qualificação
Injection Volume:	100,00 ul	Channel Name:	2487Channel 1
Run Time:	3,0 Minutes	Proc. Chnl. Descr.:	2487Channel 1
Date Acquired:	22/09/2021 13:08:19 BRT		
Date Processed:	22/09/2021 13:18:35 BRT		

Auto-Scaled Chromatogram



Peak Results

	Name	RT	Area	Height	Amount	Units
1	Cafeina	1,670	5434841	854119		