

Data File C:\CHEM32\1\DATA\QUAL_2019\EDUV000006.D
Sample Name: EDUV

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)
Acq. Instrument : HPLC 1100                       Location : Vial 1
Injection Date  : 15/10/2019 10:38:10
                                                  Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 10:36:32 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 10:43:29 by Rodolfo (QPack)
                  (modified after loading)
Sample Info     : Teste de Exatidão do Comprimento de Onda
```



Data File C:\CHEM32\1\DATA\QUAL_2019\EDUV000006.D
Sample Name: EDUV




```
=====================================================================
                      External Standard Report
=====================================================================

Sorted By             :      Signal
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Sample Amount:                :     1.00000  [ng/ul]   (not used in calc.)
Use Multiplier & Dilution Factor with ISTDs
=====================================================================
=====================================================================
                         Area Percent Report
=====================================================================

Sorted By             :      Signal
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Sample Amount:                :     1.00000  [ng/ul]   (not used in calc.)
Use Multiplier & Dilution Factor with ISTDs


Signal 1: MWD1 A, Sig=203,16 Ref=360,100

Peak RetTime Type  Width     Area      Height     Area
  #   [min]        [min]   [mAU*s]     [mAU]       %
----|-------|----|-------|----------|----------|--------|
   1   3.082 BBA   0.3024  716.99982   36.30711 100.0000

Totals :                   716.99982   36.30711



Signal 2: MWD1 B, Sig=204,16 Ref=360,100

Peak RetTime Type  Width     Area      Height     Area
  #   [min]        [min]   [mAU*s]     [mAU]       %
----|-------|----|-------|----------|----------|--------|
   1   3.081 BBA   0.3016  722.74561   36.73248 100.0000

Totals :                   722.74561   36.73248
```

Data File C:\CHEM32\1\DATA\QUAL_2019\EDUV000006.D
Sample Name: EDUV

Signal 3: MWD1 C, Sig=205,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	3.082	BBA	0.3008	722.25916	36.83632	100.0000
Totals :				722.25916	36.83632	

Signal 4: MWD1 D, Sig=206,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	3.082	BBA	0.3020	716.48438	36.66041	100.0000
Totals :				716.48438	36.66041	

Signal 5: MWD1 E, Sig=207,16 Ref=360,100


Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	3.082	BBA	0.3015	706.29498	36.22368	100.0000
Totals :				706.29498	36.22368	

==========


*** End of Report ***

Data File C:\CHEM32\1\DATA\QUAL_2019\EDUV000007.D
Sample Name: EDUV

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)
Acq. Instrument : HPLC 1100                       Location : Vial 1
Injection Date  : 15/10/2019 10:47:05
                                                  Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 10:45:19 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 10:43:29 by Rodolfo (QPack)
                  (modified after loading)
Sample Info     : Teste de Exatidão do Comprimento de Onda
```



Data File C:\CHEM32\1\DATA\QUAL_2019\EDUV000007.D
Sample Name: EDUV




```
=====================================================================
                      External Standard Report
=====================================================================

Sorted By             :      Signal
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Sample Amount:                :     1.00000  [ng/ul]   (not used in calc.)
Use Multiplier & Dilution Factor with ISTDs
=====================================================================
=====================================================================
                         Area Percent Report
=====================================================================

Sorted By             :      Signal
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Sample Amount:                :     1.00000  [ng/ul]   (not used in calc.)
Use Multiplier & Dilution Factor with ISTDs


Signal 1: MWD1 A, Sig=243,16 Ref=360,100

Peak RetTime Type  Width     Area      Height     Area
  #   [min]        [min]   [mAU*s]     [mAU]       %
----|-------|----|-------|----------|----------|--------|
   1   3.090 BB    0.2963   92.95068    4.83399 100.0000

Totals :                    92.95068    4.83399



Signal 2: MWD1 B, Sig=244,16 Ref=360,100

Peak RetTime Type  Width     Area      Height     Area
  #   [min]        [min]   [mAU*s]     [mAU]       %
----|-------|----|-------|----------|----------|--------|
   1   3.090 BB    0.2962   91.64304    4.76941 100.0000

Totals :                    91.64304    4.76941
```

Data File C:\CHEM32\1\DATA\QUAL_2019\EDUV000007.D
Sample Name: EDUV

Signal 3: MWD1 C, Sig=245,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	3.090	BB	0.2961	91.30688	4.75383	100.0000
Totals :				91.30688	4.75383	

Signal 4: MWD1 D, Sig=246,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	3.090	BB	0.2959	91.90402	4.78804	100.0000
Totals :				91.90402	4.78804	

Signal 5: MWD1 E, Sig=247,16 Ref=360,100


Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	3.090	BB	0.2957	93.50147	4.87668	100.0000
Totals :				93.50147	4.87668	

==========


*** End of Report ***

Data File C:\CHEM32\1\DATA\QUAL_2019\EDUV000008.D
Sample Name: EDUV

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)
Acq. Instrument : HPLC 1100                       Location : Vial 1
Injection Date  : 15/10/2019 10:54:27
                                                 Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 10:52:44 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 10:43:29 by Rodolfo (QPack)
                  (modified after loading)
Sample Info     : Teste de Exatidão do Comprimento de Onda
```



Data File C:\CHEM32\1\DATA\QUAL_2019\EDUV000008.D
Sample Name: EDUV




```
=====================================================================
                      External Standard Report
=====================================================================

Sorted By             :      Signal
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Sample Amount:                :     1.00000  [ng/ul]   (not used in calc.)
Use Multiplier & Dilution Factor with ISTDs
=====================================================================
=====================================================================
                         Area Percent Report
=====================================================================

Sorted By             :      Signal
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Sample Amount:                :     1.00000  [ng/ul]   (not used in calc.)
Use Multiplier & Dilution Factor with ISTDs


Signal 1: MWD1 A, Sig=271,16 Ref=360,100

Peak RetTime Type  Width     Area      Height     Area
  #   [min]        [min]   [mAU*s]     [mAU]       %
----|-------|----|-------|----------|----------|--------|
   1   3.093 BB    0.2918  264.92413   13.93697 100.0000

Totals :                   264.92413   13.93697



Signal 2: MWD1 B, Sig=272,16 Ref=360,100

Peak RetTime Type  Width     Area      Height     Area
  #   [min]        [min]   [mAU*s]     [mAU]       %
----|-------|----|-------|----------|----------|--------|
   1   3.093 BB    0.2918  266.58688   14.02545 100.0000

Totals :                   266.58688   14.02545
```

Data File C:\CHEM32\1\DATA\QUAL_2019\EDUV000008.D
Sample Name: EDUV

Signal 3: MWD1 C, Sig=273,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	3.093	BB	0.2917	266.94843	14.04532	100.0000
Totals :				266.94843	14.04532	

Signal 4: MWD1 D, Sig=274,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	3.094	BB	0.2917	265.90967	13.99138	100.0000
Totals :				265.90967	13.99138	

Signal 5: MWD1 E, Sig=275,16 Ref=360,100


Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Height [mAU]	Area %
1	3.094	BB	0.2917	263.46521	13.86353	100.0000
Totals :				263.46521	13.86353	

==========

*** End of Report ***

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000001.D
Sample Name: Branco 1

```
   ======================================================================
   Acq. Operator   : Rodolfo (QPack)              Seq. Line :   1
   Acq. Instrument : HPLC 1100                    Location : Vial 2
   Injection Date  : 15/10/2019 11:03:11               Inj :   1
                                                 Inj Volume : 10.0 µl
   Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\QUALIFICAÇÃO.M
   Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
   Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
   Last changed    : 15/10/2019 10:43:29 by Rodolfo (QPack)
                     (modified after loading)
```




```
   ======================================================================
                        External Standard Report
   ======================================================================

   Sorted By             :      Signal
   Multiplier:                   :      1.0000
   Dilution:                     :      1.0000
   Use Multiplier & Dilution Factor with ISTDs
   ======================================================================
   ======================================================================
                           Area Percent Report
   ======================================================================


   Sorted By             :      Signal
   Multiplier:                   :      1.0000
   Dilution:                     :      1.0000
   Use Multiplier & Dilution Factor with ISTDs

   No peaks found

   ======================================================================
                           *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000002.D
Sample Name: PCIN 1

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   2
Acq. Instrument : HPLC 1100                    Location : Vial 3
Injection Date  : 15/10/2019 11:08:39               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:17:03 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.091 BBA     799.35681 1.25101e-1  100.00000    Cafeína

Totals :                               100.00000


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000003.D
Sample Name: PCIN 2

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   3
Acq. Instrument : HPLC 1100                    Location : Vial 3
Injection Date  : 15/10/2019 11:14:03               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:17:03 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.094 BB      798.26764 1.25101e-1   99.86374    Cafeína

Totals :                                99.86374


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000004.D
Sample Name: PCIN 3

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   4
Acq. Instrument : HPLC 1100                    Location : Vial 3
Injection Date  : 15/10/2019 11:19:30               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:17:03 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.100 BBA     800.31146 1.25101e-1  100.11943    Cafeína

Totals :                               100.11943


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000005.D
Sample Name: Branco 2

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   5
Acq. Instrument : HPLC 1100                    Location : Vial 2
Injection Date  : 15/10/2019 11:24:56               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:17:03 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                      External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100

RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.091            -          -          -         Cafeína

Totals :                                0.00000


1 Warnings or Errors :

Warning : Calibrated compound(s) not found

==========================================================================
```

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000005.D
Sample Name: Branco 2

==========

Area Percent Report

==========

Sorted By : Signal
Calib. Data Modified : Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier: : 1.0000
Dilution: : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Area %	Name
1	3.091		0.0000	0.00000	0.0000	Cafeína

Totals : 0.00000

1 Warnings or Errors :


Warning : Calibrated compound(s) not found

==========

*** End of Report ***

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000006.D
Sample Name: PCIN 4

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)               Seq. Line :   6
Acq. Instrument : HPLC 1100                     Location : Vial 3
Injection Date  : 15/10/2019 11:30:25                Inj :   1
                                               Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:17:03 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.106 BBA     801.04510 1.25101e-1  100.21121    Cafeína

Totals :                               100.21121


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000007.D
Sample Name: PCIN 5

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)               Seq. Line :   7
Acq. Instrument : HPLC 1100                     Location : Vial 3
Injection Date  : 15/10/2019 11:35:52                Inj :   1
                                               Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:17:03 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.112 BBA     801.50537 1.25101e-1  100.26879    Cafeína

Totals :                               100.26879


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000008.D
Sample Name: PCIN 6

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)               Seq. Line :   8
Acq. Instrument : HPLC 1100                     Location : Vial 3
Injection Date  : 15/10/2019 11:41:14                Inj :   1
                                               Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:17:03 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.114 BBA     800.72528 1.25101e-1  100.17120    Cafeína

Totals :                               100.17120


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000009.D
Sample Name: Branco 3

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   9
Acq. Instrument : HPLC 1100                    Location : Vial 2
Injection Date  : 15/10/2019 11:46:41               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:17:03 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100

RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.091            -          -          -         Cafeína

Totals :                                0.00000


1 Warnings or Errors :

Warning : Calibrated compound(s) not found

==========================================================================
```

Data File C:\CHEM32\1\DATA\QUAL_2019\PCIN 2019-10-15 11-01-58\PCIN0000009.D
Sample Name: Branco 3

==========

Area Percent Report

==========

Sorted By : Signal
Calib. Data Modified : Tuesday, 15 15e October 15e 2019 11:16:48
Multiplier: : 1.0000
Dilution: : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Area %	Name
1	3.091		0.0000	0.00000	0.0000	Cafeína

Totals : 0.00000

1 Warnings or Errors :


Warning : Calibrated compound(s) not found

==========

*** End of Report ***

Data File C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\LDUV0000001.D
Sample Name: LDUV 1

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)               Seq. Line :   1
Acq. Instrument : HPLC 1100                     Location : Vial 4
Injection Date  : 15/10/2019 12:00:48                Inj :   1
                                               Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100

RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.091            -          -          -         Cafeína

Totals :                                0.00000


1 Warnings or Errors :

Warning : Calibrated compound(s) not found

==========================================================================
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\LDUV0000001.D
Sample Name: LDUV 1

==========

Area Percent Report

==========

Sorted By : Signal
Calib. Data Modified : 15/10/2019 12:31:08
Multiplier: : 1.0000
Dilution: : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Area %	Name
1	3.091		0.0000	0.00000	0.0000	Cafeína

Totals : 0.00000

1 Warnings or Errors :


Warning : Calibrated compound(s) not found

==========

*** End of Report ***

Data File C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\LDUV0000002.D
Sample Name: LDUV 2

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   2
Acq. Instrument : HPLC 1100                    Location : Vial 5
Injection Date  : 15/10/2019 12:08:56               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.117 BB      276.94586 1.25101e-1   34.64609    Cafeína

Totals :                                34.64609


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\LDUV0000003.D
Sample Name: LDUV 3

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   3
Acq. Instrument : HPLC 1100                    Location : Vial 6
Injection Date  : 15/10/2019 12:14:21               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.114 BBA     521.71210 1.25101e-1   65.26649    Cafeína

Totals :                                65.26649


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\LDUV0000004.D
Sample Name: LDUV 4

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   4
Acq. Instrument : HPLC 1100                    Location : Vial 7
Injection Date  : 15/10/2019 12:19:43               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.119 BBA     785.58710 1.25101e-1   98.27740    Cafeína

Totals :                                98.27740


==========================================================================
                         *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\LDUV0000005.D
Sample Name: LDUV 5

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   5
Acq. Instrument : HPLC 1100                    Location : Vial 8
Injection Date  : 15/10/2019 12:25:12               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDUV 2019-10-15 11-59-37\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 11:01:56 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.126 BBA    1030.61328 1.25101e-1  128.93032    Cafeína

Totals :                               128.93032


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\LIIN0000001.D
Sample Name: LIIN 1

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   1
Acq. Instrument : HPLC 1100                    Location : Vial 9
Injection Date  : 15/10/2019 12:39:37               Inj :   1
                                              Inj Volume : 10.0 µl
Different Inj Volume from Sequence !     Actual Inj Volume : 0.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:38:24 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                      External Standard Report
--------------------------------------------------------------------------

Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100

RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.091            -          -          -         Cafeína

Totals :                                0.00000


1 Warnings or Errors :

Warning : Calibrated compound(s) not found

==========================================================================
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\LIIN0000001.D
Sample Name: LIIN 1

==========

Area Percent Report

==========

Sorted By : Signal
Calib. Data Modified : 15/10/2019 12:31:08
Multiplier: : 1.0000
Dilution: : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100

Peak #	RetTime [min]	Type	Width [min]	Area [mAU*s]	Area %	Name
1	3.091		0.0000	0.00000	0.0000	Cafeína

Totals : 0.00000

1 Warnings or Errors :


Warning : Calibrated compound(s) not found

==========

*** End of Report ***

Data File C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\LIIN0000002.D
Sample Name: LIIN 2

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   2
Acq. Instrument : HPLC 1100                    Location : Vial 9
Injection Date  : 15/10/2019 12:44:55               Inj :   1
                                              Inj Volume : 10.0 µl
Different Inj Volume from Sequence !     Actual Inj Volume : 5.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:38:24 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                      External Standard Report
--------------------------------------------------------------------------

Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.127 BBA     397.05038 1.25101e-1   49.67123    Cafeína

Totals :                               49.67123


==========================================================================
                         *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\LIIN0000003.D
Sample Name: LIIN 3

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   3
Acq. Instrument : HPLC 1100                    Location : Vial 9
Injection Date  : 15/10/2019 12:50:18               Inj :   1
                                              Inj Volume : 10.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:38:24 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.136 BBA     792.28802 1.25101e-1   99.11569    Cafeína

Totals :                                99.11569


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\LIIN0000004.D
Sample Name: LIIN 4

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   4
Acq. Instrument : HPLC 1100                    Location : Vial 9
Injection Date  : 15/10/2019 12:55:54               Inj :   1
                                              Inj Volume : 10.0 µl
Different Inj Volume from Sequence !     Actual Inj Volume : 15.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:38:24 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                      External Standard Report
--------------------------------------------------------------------------

Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.133 BBA    1186.85718 1.25101e-1  148.47652    Cafeína

Totals :                              148.47652


==========================================================================
                         *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\LIIN0000005.D
Sample Name: LIIN 5

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   5
Acq. Instrument : HPLC 1100                    Location : Vial 9
Injection Date  : 15/10/2019 13:01:20               Inj :   1
                                              Inj Volume : 10.0 µl
Different Inj Volume from Sequence !     Actual Inj Volume : 20.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LIIN 2019-10-15 12-38-26\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:38:24 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 12:35:28 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                      External Standard Report
--------------------------------------------------------------------------

Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 12:31:08
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: MWD1 A, Sig=273,16 Ref=360,100

RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]          [mAU*s]               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  3.144 BBA    1578.97107 1.25101e-1  197.53020    Cafeína


Totals :                              197.53020


==========================================================================
                         *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\EDFL000002.D
Sample Name: EDFL

==========

Acq. Operator : Rodolfo (QPack)
Acq. Instrument : HPLC 1100 Location : Vial 11
Injection Date : 15/10/2019 14:14:09
Inj Volume : 10.0 µl
Acq. Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed : 15/10/2019 14:12:17 by Rodolfo (QPack)
(modified after loading)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed : 15/10/2019 14:02:32 by Rodolfo (QPack)
(modified after loading)
Sample Info : Teste de Exatidão do Comprimento de Onda do FLD



Data File C:\CHEM32\1\DATA\QUAL_2019\EDFL000002.D
Sample Name: EDFL

==========
External Standard Report
==========

Sorted By : Signal
Multiplier: : 1.0000
Dilution: : 1.0000
Sample Amount: : 1.00000 [ng/ul] (not used in calc.)
Use Multiplier & Dilution Factor with ISTDs

==========
==========
Area Percent Report
==========

Sorted By : Signal
Multiplier: : 1.0000
Dilution: : 1.0000
Sample Amount: : 1.00000 [ng/ul] (not used in calc.)
Use Multiplier & Dilution Factor with ISTDs

Signal 1: FLD1 A, Ex=348, Em=445

Peak #	RetTime [min]	Type	Width [min]	LU	Area *s	Height [LU]	Area %
1	0.835	BB	0.3249		9.63773	4.63043e-1	100.0000
Totals :					9.63773	4.63043e-1	

Signal 2: FLD1 B, Ex=348, Em=446

Peak #	RetTime [min]	Type	Width [min]	LU	Area *s	Height [LU]	Area %
1	0.846	BB	0.3259		9.24670	4.50483e-1	100.0000
Totals :					9.24670	4.50483e-1	

Signal 3: FLD1 C, Ex=348, Em=447

Peak #	RetTime [min]	Type	Width [min]	LU	Area *s	Height [LU]	Area %
1	0.838	BB	0.3092		9.21248	4.47687e-1	100.0000
Totals :					9.21248	4.47687e-1	

Signal 4: FLD1 D, Ex=348, Em=448

Peak #	RetTime [min]	Type	Width [min]	LU	Area *s	Height [LU]	Area %
1	0.840	BB	0.3293		9.18866	4.37383e-1	100.0000

Data File C:\CHEM32\1\DATA\QUAL_2019\EDFL000002.D
Sample Name: EDFL


Totals :	9.18866	4.37383e-1

==========

*** End of Report ***

Data File C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\LDFL0000001.D
Sample Name: LDFL 1

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   1
Acq. Instrument : HPLC 1100                    Location : Vial 2
Injection Date  : 15/10/2019 15:43:18               Inj :   1
                                              Inj Volume : 20.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 15:40:45 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 16:07:54 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
  Sorted By             :      Signal
  Calib. Data Modified  :      15/10/2019 16:07:44
  Multiplier:                   :      1.0000
  Dilution:                     :      1.0000
  Use Multiplier & Dilution Factor with ISTDs

  Signal 1: FLD1 A, Ex=348, Em=447

  RetTime  Type     Area     Amt/Area    Amount   Grp   Name
   [min]         LU    *s               [ng/ul]
  -------|------|----------|----------|----------|--|------------------
    0.860            -          -          -         Quinina

  Totals :                                0.00000


  1 Warnings or Errors :

  Warning : Calibrated compound(s) not found

  ==========================================================================
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\LDFL0000001.D
Sample Name: LDFL 1

==========

Area Percent Report

==========

Sorted By : Signal
Calib. Data Modified : 15/10/2019 16:07:44
Multiplier: : 1.0000
Dilution: : 1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: FLD1 A, Ex=348, Em=447

Peak #	RetTime [min]	Type	Width [min]	Area LU *s	Area %	Name
1	0.860		0.0000	0.00000	0.0000	Quinina

Totals : 0.00000

1 Warnings or Errors :


Warning : Calibrated compound(s) not found

==========

*** End of Report ***

Data File C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\LDFL0000002.D
Sample Name: LDFL 2

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   2
Acq. Instrument : HPLC 1100                    Location : Vial 13
Injection Date  : 15/10/2019 15:47:56               Inj :   1
                                              Inj Volume : 20.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 15:40:45 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 16:07:54 by Rodolfo (QPack)
                  (modified after loading)
```




```
==========================================================================
                        External Standard Report
==========================================================================
Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 16:07:44
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: FLD1 A, Ex=348, Em=447


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]         LU    *s               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  0.860 BB       77.67221 1.31603e-2    1.02219    Quinina

Totals :                                1.02219


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\LDFL0000003.D
Sample Name: LDFL 3

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)               Seq. Line :   3
Acq. Instrument : HPLC 1100                     Location : Vial 13
Injection Date  : 15/10/2019 15:52:43                Inj :   1
                                               Inj Volume : 20.0 µl
Different Inj Volume from Sequence !     Actual Inj Volume : 40.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 15:40:45 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 16:07:54 by Rodolfo (QPack)
                  (modified after loading)
Additional Info : Peak(s) manually integrated
```




```
==========================================================================
                     External Standard Report
==========================================================================

Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 16:07:44
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: FLD1 A, Ex=348, Em=447


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]         LU    *s               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  0.868 MM      156.05469 1.31603e-2    2.05373    Quinina

Totals :                                2.05373


==========================================================================
                          *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\LDFL0000004.D
Sample Name: LDFL 4

```
==================================================================
Acq. Operator   : Rodolfo (QPack)              Seq. Line :   4
Acq. Instrument : HPLC 1100                    Location : Vial 13
Injection Date  : 15/10/2019 15:57:51               Inj :   1
                                              Inj Volume : 20.0 µl
Different Inj Volume from Sequence !     Actual Inj Volume : 60.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 15:40:45 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 16:07:54 by Rodolfo (QPack)
                  (modified after loading)
```




```
==================================================================
                     External Standard Report
------------------------------------------------------------------

Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 16:07:44
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: FLD1 A, Ex=348, Em=447


RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]         LU    *s               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  0.868 BB      236.79686 1.31603e-2    3.11633    Quinina

Totals :                                3.11633


==================================================================
                        *** End of Report ***
```

Data File C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\LDFL0000005.D
Sample Name: LDFL 5

```
==========================================================================
Acq. Operator   : Rodolfo (QPack)               Seq. Line :   5
Acq. Instrument : HPLC 1100                     Location : Vial 13
Injection Date  : 15/10/2019 16:03:24                Inj :   1
                                               Inj Volume : 20.0 µl
Different Inj Volume from Sequence !     Actual Inj Volume : 80.0 µl
Acq. Method     : C:\CHEM32\1\DATA\QUAL_2019\LDFL 2019-10-15 15-41-18\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 15:40:45 by Rodolfo (QPack)
Analysis Method : C:\CHEM32\1\METHODS\QUALIFICAÇÃO.M
Last changed    : 15/10/2019 16:07:54 by Rodolfo (QPack)
                  (modified after loading)
Additional Info : Peak(s) manually integrated
```




```
==========================================================================
                     External Standard Report
==========================================================================

Sorted By             :      Signal
Calib. Data Modified  :      15/10/2019 16:07:44
Multiplier:                   :      1.0000
Dilution:                     :      1.0000
Use Multiplier & Dilution Factor with ISTDs

Signal 1: FLD1 A, Ex=348, Em=447

RetTime  Type     Area     Amt/Area    Amount   Grp   Name
 [min]         LU    *s               [ng/ul]
-------|------|----------|----------|----------|--|------------------
  0.877 MM      321.50174 1.31603e-2    4.23107    Quinina

Totals :                                4.23107


==========================================================================
                          *** End of Report ***
```