Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\EDUV 001.sirslt\EDUV 001.dx		
Sample name:	EDUV 001		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C1
Injection date:	2024-05-28 14:22:02-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Cafeína.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:28:37-03:00		

RT:

1.009


Compound Name:

Cafeína



Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\RELB 001.sirslt\RELB 001.dx		
Sample name:	RELB 001		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C8
Injection date:	2024-05-28 14:28:18-03:00	**Injection:**	1 of 1
Acq. method:	Sinal Ruído.amx	**Injection volume:**	20.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:56:13-03:00		



Sample Name	Name	RT	Area	Height
RELB 001	Cafeína	0.972	811.3628	300.755

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\RELB 001.sirslt\RELB 001.dx		
Sample name:	RELB 001		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C8
Injection date:	2024-05-28 14:28:18-03:00	**Injection:**	1 of 1
Acq. method:	Sinal Ruído.amx	**Injection volume:**	20.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:56:13-03:00		



Sample Name	Name	RT	Area	Height
RELB 001	Cafeína	0.972	811.3628	300.755

Single Injection Report

Data file:	C:\CDSProjects\Qualificação\Results\2024\RELB 001.sirslt\RELB 001.dx		
Sample name:	RELB 001		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C8
Injection date:	2024-05-28 14:28:18-03:00	**Injection:**	1 of 1
Acq. method:	Sinal Ruído.amx	**Injection volume:**	20.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:56:13-03:00		



Sample Name	Name	RT	Area	Height
RELB 001	Cafeína	0.972	811.3628	300.755

Certificado de Qualificação

Certificado Nº:	Início da Qualificação	Página 1 de 2

Resultados obtidos no teste estabilidade da linha de base

Estabilidade da Linha de Base	Diferença 20 min	Diferença 60 min
VALOR MÁXIMO:	**-0,2000**	**2,1000**
VALOR MÍNIMO:	**-0,9000**	
Resultado:	**0,7000**	

Resultado e critérios de aceitação:

Parâmetros	Critério de Aceitação	Resultados Obtidos	Resultado
Estabilidade da Linha de Base	≤ 5,00 mAU\h	**2,100 mAU\h**	***PASSOU***

Rua Ema Gazzi Magnusson, 138 - Comercial Vitória Martini - Indaiatuba - SP - CEP:13347-630
www.labprovider.com.br - Fone: 19 2660-1110

Certificado de Qualificação

Certificado Nº:	Início da Qualificação:	Página 1 de 2

Resultados obtidos no teste de Sinal/Ruído

	ALTURA DO PICO (mAU)
Resultado:	**301**

Resultado:	Ruído (mAU)
VALOR MÁXIMO:	**-0,8250**
VALOR MÍNIMO:	**-0,8650**
	Ruído
	0,0400


SINAL/Ruído
7519

Resultado e critérios de aceitação:

Parâmetros	Critério de Aceitação	Resultados Obtidos	Resultado
Sinal/Ruído	≥ 3000,00 S/R	**7519 S/R**	***PASSOU***
Ruído	≤ 0,05 mAU	**0,040 mAU**	***PASSOU***

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\PCIN 002.rslt\Branco PCIN 012024-05-28 15-24-06-03-00.dx		
Sample name:	Branco PCIN 01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C2
Injection date:	2024-05-28 15:24:58-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height

Single Injection Report

Data file:	C:\CDSProjects\Qualificação\Results\2024\PCIN 002.rslt\PCIN 012024-05-28 15-27-02-03-00.dx		
Sample name:	PCIN 01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C1
Injection date:	2024-05-28 15:27:56-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 15:56:54-03:00		




Sample Name	Name	RT	Area	Height
PCIN 01	Cafeína	1.135	993.3913	320.053

Single Injection Report

Agilent

Data file:	C:\CDSProjects\Qualificação\Results\2024\PCIN 002.rslt\PCIN 022024-05-28 15-30-01-03-00.dx		
Sample name:	PCIN 02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C1
Injection date:	2024-05-28 15:30:55-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 15:56:54-03:00		




Sample Name	Name	RT	Area	Height
PCIN 02	Cafeína	1.110	987.4270	311.500

Single Injection Report

Agilent


Data file:	C:\CDSProjects\Qualificação\Results\2024\PCIN 002.rslt\PCIN 032024-05-28 15-32-59-03-00.dx		
Sample name:	PCIN 03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C1
Injection date:	2024-05-28 15:33:53-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 16:04:27-03:00		



Sample Name	Name	RT	Area	Height
PCIN 03	Cafeína	1.109	989.7851	307.377

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\PCIN 002.rslt\Branco PCIN 022024-05-28 15-35-57-03-00.dx		
Sample name:	Branco PCIN 02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C2
Injection date:	2024-05-28 15:36:48-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 15:56:54-03:00		



Sample Name	Name	RT	Area	Height

Single Injection Report

Data file:	C:\CDSProjects\Qualificação\Results\2024\PCIN 002.rslt\PCIN 042024-05-28 15-38-52-03-00.dx		
Sample name:	PCIN 04		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C1
Injection date:	2024-05-28 15:39:46-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 15:56:54-03:00		




Sample Name	Name	RT	Area	Height
PCIN 04	Cafeína	1.081	988.9455	322.045

Single Injection Report

Agilent

Data file:	C:\CDSProjects\Qualificação\Results\2024\PCIN 002.rslt\PCIN 052024-05-28 15-41-51-03-00.dx		
Sample name:	PCIN 05		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C1
Injection date:	2024-05-28 15:42:45-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 16:06:15-03:00		




Sample Name	Name	RT	Area	Height
PCIN 05	Cafeína	1.136	996.5153	321.145

Single Injection Report

Agilent


Data file:	C:\CDSProjects\Qualificação\Results\2024\PCIN 002.rslt\PCIN 062024-05-28 15-44-48-03-00.dx		
Sample name:	PCIN 06		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C1
Injection date:	2024-05-28 15:45:43-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 15:56:54-03:00		



Sample Name	Name	RT	Area	Height
PCIN 06	Cafeína	1.108	985.5686	320.352

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\PCIN 002.rslt\Branco PCIN 032024-05-28 15-47-47-03-00.dx		
Sample name:	Branco PCIN 03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C2
Injection date:	2024-05-28 15:48:40-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 15:56:54-03:00		



Sample Name	Name	RT	Area	Height

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\LDUV 003.rslt\LDUV 012024-05-28 16-23-44-03-00.dx		
Sample name:	LDUV 01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C2
Injection date:	2024-05-28 16:24:40-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height

Single Injection Report

Data file:	C:\CDSProjects\Qualificação\Results\2024\LDUV 003.rslt\LDUV 022024-05-28 16-26-44-03-00.dx		
Sample name:	LDUV 02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C3
Injection date:	2024-05-28 16:27:34-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		




Sample Name	Name	RT	Area	Height
LDUV 02	Cafeína	1.041	459.6111	136.541

Single Injection Report

Agilent


Data file:	C:\CDSProjects\Qualificação\Results\2024\LDUV 003.rslt\LDUV 032024-05-28 16-29-38-03-00.dx		
Sample name:	LDUV 03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C4
Injection date:	2024-05-28 16:30:30-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height
LDUV 03	Cafeína	1.064	976.5877	285.908

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\LDUV 003.rslt\LDUV 042024-05-28 16-32-34-03-00.dx		
Sample name:	LDUV 04		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C5
Injection date:	2024-05-28 16:33:25-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height
LDUV 04	Cafeína	1.026	1417.6145	429.593

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\LDUV 003.rslt\LDUV 052024-05-28 16-35-31-03-00.dx		
Sample name:	LDUV 05		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C6
Injection date:	2024-05-28 16:36:24-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height
LDUV 05	Cafeína	1.042	1908.9037	585.769

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\LIIN 002.rslt\LIIN 012024-05-28 16-38-39-03-00.dx		
Sample name:	LIIN 01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C7
Injection date:	2024-05-28 16:39:27-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	0.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\LIIN 002.rslt\LIIN 022024-05-28 16-41-32-03-00.dx		
Sample name:	LIIN 02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C7
Injection date:	2024-05-28 16:42:20-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	5.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height
LIIN 02	Cafeína	1.048	751.9960	216.647

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\LIIN 002.rslt\LIIN 032024-05-28 16-44-23-03-00.dx		
Sample name:	LIIN 03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C7
Injection date:	2024-05-28 16:45:15-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height
LIIN 03	Cafeína	1.048	1498.1651	451.711

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\LIIN 002.rslt\LIIN 042024-05-28 16-47-20-03-00.dx		
Sample name:	LIIN 04		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C7
Injection date:	2024-05-28 16:48:13-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	15.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height
LIIN 04	Cafeína	1.062	2254.4742	686.062

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\LIIN 002.rslt\LIIN 052024-05-28 16-50-18-03-00.dx		
Sample name:	LIIN 05		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C7
Injection date:	2024-05-28 16:51:15-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	20.000
Analysis method:	Cafeína 2.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-28 14:55:15-03:00		



Sample Name	Name	RT	Area	Height
LIIN 05	Cafeína	1.072	3027.5095	920.994

Single Injection Report


Data file:	C:\CDSProjects\Qualificação\Results\2024\GRAD AB 002.sirslt\GRAD AB 002.dx		
Sample name:	GRAD AB 002		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C2
Injection date:	2024-05-29 09:21:22-03:00	**Injection:**	1 of 1
Acq. method:	Gradiente AB.amx	**Injection volume:**	0.000
Analysis method:	*GC_LC Area Percent_DefaultMethod.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-29 09:44:19-03:00		



Sample Name	Name	RT	Area	Height
GRAD AB 002	20%	3.715	4816.0279	32.961
GRAD AB 002	40%	6.955	10696.7053	66.091
GRAD AB 002	60%	9.983	16743.3824	99.229
GRAD AB 002	80%	12.748	24172.8012	132.775
GRAD AB 002	100%	16.317	33054.9796	166.232

Single Injection Report

Data file:	C:\CDSProjects\Qualificação\Results\2024\GRAD CD 002.sirslt\GRAD CD 002.dx		
Sample name:	GRAD CD 002		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	HPLC	**Location:**	P2-C2
Injection date:	2024-05-29 09:53:50-03:00	**Injection:**	1 of 1
Acq. method:	Gradiente CD.amx	**Injection volume:**	0.000
Analysis method:	*Gradiente.pmx	**Acq. operator:**	SYSTEM
Last changed:	2024-05-29 10:18:43-03:00		



Sample Name	Name	RT	Area	Height
GRAD CD 002	20%	4.121	5059.9076	32.878
GRAD CD 002	40%	6.926	10926.6259	65.621
GRAD CD 002	60%	10.119	17116.2321	98.644
GRAD CD 002	80%	12.752	21533.9883	131.480
GRAD CD 002	100%	16.236	34258.3936	165.572