Inspection Parameters - Repetibilidade OR.orvp

Creator	JASCO
Workgroup	Suporte Tecnico
Division	Suporte Técnico
Company	EQUIPLEX
Creation Date	30/05/2023 11:00
Instrument Name	POLARIMETRO
Model Name	POLARIMETRO
Serial No.	A098461232
Polarizer	Dichrom
Faraday Cell	Flint Glass
Light Source	Na
Monitor wavelength	589 nm
D.I.T.	5 sec
No. of cycle	1
Cycle interval	0 sec
Aperture(S)	8.0mm
Aperture(L)	Auto
Mode	Optical Rotation
Inspection Method	Test

Repeatability of O.R.

Acceptable criterion	Less than or equal to 0.01 degree(s)
Measurements	10
Test sample	Rotatory plate

Creation	QPackSA
Review	QPackSA
Approval	

Inspection Sheet

Date	quinta-feira, 28 de maio de 2026
Model Name	Polarimetro
Serial No.	A098461232
Repeatability of O.R.	Pass/Fail : Pass
Rotatory plate	Criteria Less than 0.0100 degree(s) Minimum 8.6981 degree(s), Maximum 8.7025 degree(s), (Max-Min) 0.0044 degree(s) 1: 8.6996 deg, 2: 8.6981 deg, 3: 8.6999 deg, 4: 8.7005 deg, 5: 8.6994 deg 6: 8.6998 deg, 7: 8.7025 deg, 8: 8.7018 deg, 9: 8.7018 deg, 10: 8.7010 deg

Unsigned

Creation
Review
Approval