Inspection Parameters - Repetibilidade ZERO.orvp

| | |
|---|---|
| Creator | JASCO |
| Workgroup | Suporte Tecnico |
| Division | Suporte Técnico |
| Company | EQUIPLEX |
| Creation Date | 30/05/2023 11:18 |
| Instrument Name | POLARIMETRO |
| Model Name | POLARIMETRO |
| Serial No. | A098461232 |
| Polarizer | Dichrom |
| Faraday Cell | Flint Glass |
| Light Source | Na |
| Monitor wavelength | 589 nm |
| D.I.T. | 5 sec |
| No. of cycle | 1 |
| Cycle interval | 0 sec |
| Aperture(S) | 8.0mm |
| Aperture(L) | Auto |
| Mode | Optical Rotation |
| Inspection Method | Test |
| Repeatability of Zero | |
| Acceptable criterion | +/- 0.01 degree(s) |
| Measurements | 10 |

| | |
|---|---|
| Creation | QPackSA |
| Review | QPackSA |
| Approval | |

Inspection Sheet

| | |
|---|---|
| Date | quinta-feira, 28 de maio de 2026 |
| Model Name | Polarimetro |
| Serial No. | A098461232 |

Repeatability of Zero

Pass/Fail : Pass

Criteria Less than 0.0100 degree(s)
Minimum -0.0004 degree(s), Maximum 0.0015 degree(s), (Max-Min) 0.0019 degree(s)
1: -0.0002 deg, 2: 0.0012 deg, 3: 0.0015 deg, 4: 0.0010 deg, 5: -0.0004 deg
6: 0.0000 deg, 7: 0.0013 deg, 8: 0.0011 deg, 9: 0.0002 deg, 10: 0.0003 deg

Unsigned

Creation
Review
Approval