
Peak Find - EDUV HPL 25.JWS



[Comments]

User	Administrator
Division	---
Company	Jasco Br
Creation date	10/29/2019 14:41
Date modified	10/29/2019 14:47

[Detailed Information]

Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	200 nm
End	300 nm
Data interval	1 nm
Data points	101


[Measurement Information]

Instrument Name	MD-2010
Model Name	MD-2010
Serial No.	072260870
ROM version	V2.36

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	273	0.163182	2	244	0.043075
3	204	0.460764			

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	Branco 01
Acquisition Time	3.0 [min]




Peak Information – No data.

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	PCIN 01
Acquisition Time	3.0 [min]

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.867	911570

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	PCIN 02
Acquisition Time	3.0 [min]

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.863	911253

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	PCIN 03
Acquisition Time	3.0 [min]

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.853	910335

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	Branco 02
Acquisition Time	3.0 [min]




Peak Information – No data.

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	PCIN 04
Acquisition Time	3.0 [min]

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.837	912253

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	PCIN 05
Acquisition Time	3.0 [min]

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.823	914198

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	PCIN 06
Acquisition Time	3.0 [min]

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.863	918501

Creator *Administrator*
Reviewer
Approver


Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	0.10 [µL]
Project Name	Qualificação
Chromatogram Name	Branco 03
Acquisition Time	3.0 [min]



Peak Information – No data.

Creator *Administrator*
Reviewer
Approver


Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	LDUV01
Acquisition Time	3.0 [min]



Peak Information – No data.

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	LDUV02
Acquisition Time	3.0 [min]




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.830	288454

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	LDUV03
Acquisition Time	3.0 [min]




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.837	601258

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	LDUV04
Acquisition Time	3.0 [min]




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.837	905646

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	LDUV05
Acquisition Time	3.0 [min]




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.827	1215155

Creator *Administrator*
Reviewer
Approver


Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	0.10 [µL]
Project Name	Qualificação
Chromatogram Name	LIIN01
Acquisition Time	3.0 [min]



Peak Information – No data.

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	5.00 [µL]
Project Name	Qualificação
Chromatogram Name	LIIN02
Acquisition Time	3.0 [min]




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.817	445944

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
HPLC System Name	HPL 25
Volume	10.00 [µL]
Project Name	Qualificação
Chromatogram Name	LIIN03
Acquisition Time	3.0 [min]




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.823	907235

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	15.00 [µL]
Project Name	Qualificação
Chromatogram Name	LIIN04
Acquisition Time	3.0 [min]




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.830	1356906

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	20.00 [µL]
Project Name	Qualificação
Chromatogram Name	LIIN05
Acquisition Time	3.0 [min]




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]
1	Cafeína	5	1.837	1843534

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	0.10 [µL]
Project Name	Qualificação
Chromatogram Name	GRBO AB
Acquisition Time	19.0 [min]




Peak Information

#	Peak Name	CH	tR [min]	Height [µV]
1	20 %	5	4.027	96944
2	40 %	5	6.260	194782
3	60 %	5	9.920	288204
4	80 %	5	12.693	387455
5	100 %	5	16.213	484930

Creator
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
HPLC System Name	HPL 25
Volume	0.10 [µL]
Project Name	Qualificação
Chromatogram Name	GRBO CD
Acquisition Time	19.0 [min]



Peak Information

#	Peak Name	CH	tR [min]	Height [µV]
1	20 %	5	4.087	95402
2	40 %	5	6.420	195410
3	60 %	5	10.447	292285
4	80 %	5	12.420	392135
5	100 %	5	16.480	484278

Creator
Reviewer
Approver