
Peak Find - 01-Exatidão do comprimento de onda.jws



[Comments]

Sample name	Exatidao do Comprimento de Onda
Comment	Qualificação
User	QPackSA
Workgroup	Jasco
Division	Suporte
Company	ACG do Brasil

[Detailed Information]

Creation date	17/06/2026 13:17
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	650 nm
End	230 nm
Data interval	0.1 nm
Data points	4201

[Measurement Information]


Instrument name	V-730
Model name	V-730
Serial No.	A149461798
Measurement date	17/06/2026 13:12
Parameter file	C:\Jasco\Qualificação\2024\Exatidão Do Comprimento De Onda._2026uvsp.uvsp
Photometric mode	Abs
Measurement range	650 - 230 nm
Data interval	0.1 nm
Bandwidth	1.0 nm
Response	0.06 sec
Scan mode	Continuous
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline
No. of cycles	1

Creation	QPack Soluções Analíticas Ltda
Review	
Approval	

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	637.3	0.190158	2	536.2	0.382747
3	453.4	1.05199	4	360.7	1.01673

Peak Find - 02-Repetibilidade de comprimento de onda-01.jws



[Comments]

Sample name	Repetibilidade do Comprimento de Onda
Comment	Qualificação
User	QPackSA
Workgroup	Jasco
Division	Suporte
Company	ACG do Brasil

[Detailed Information]

Creation date	17/06/2026 13:25
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	366 nm
End	356 nm
Data interval	0.1 nm
Data points	101

[Measurement Information]

Instrument name	V-730
Model name	V-730
Serial No.	A149461798
Measurement date	17/06/2026 13:22
Parameter file	C:\Jasco\Qualificação\2024\Repetibilidade Do Comprimento De Onda._2026.uvsp
Photometric mode	Abs
Measurement range	366 - 356 nm
Data interval	0.1 nm
Bandwidth	1.0 nm
Response	0.06 sec
Scan mode	Continuous
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline


Creation	QPack Soluções Analíticas Ltda
Review	
Approval	

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	360.7	1.0151			

Creation QPack Soluções Analíticas Ltda
Review
Approval

Peak Find - 02-Repetibilidade de comprimento de onda-02.jws



[Comments]

Sample name	Repetibilidade do Comprimento de Onda
Comment	Qualificação
User	QPackSA
Workgroup	Jasco
Division	Suporte
Company	ACG do Brasil

[Detailed Information]

Creation date	17/06/2026 13:26
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	366 nm
End	356 nm
Data interval	0.1 nm
Data points	101

[Measurement Information]

Instrument name	V-730
Model name	V-730
Serial No.	A149461798
Measurement date	17/06/2026 13:23
Parameter file	C:\Jasco\Qualificação\2024\Repetibilidade Do Comprimento De Onda._2026.uvsp
Photometric mode	Abs
Measurement range	366 - 356 nm
Data interval	0.1 nm
Bandwidth	1.0 nm
Response	0.06 sec
Scan mode	Continuous
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline


Creation	QPack Soluções Analíticas Ltda
Review	
Approval	

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	360.7	1.01592			

Creation QPack Soluções Analíticas Ltda
Review
Approval

Peak Find - 02-Repetibilidade de comprimento de onda-03.jws



[Comments]

Sample name	Repetibilidade do Comprimento de Onda
Comment	Qualificação
User	QPackSA
Workgroup	Jasco
Division	Suporte
Company	ACG do Brasil

[Detailed Information]

Creation date	17/06/2026 13:26
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	366 nm
End	356 nm
Data interval	0.1 nm
Data points	101

[Measurement Information]

Instrument name	V-730
Model name	V-730
Serial No.	A149461798
Measurement date	17/06/2026 13:23
Parameter file	C:\Jasco\Qualificação\2024\Repetibilidade Do Comprimento De Onda._2026.uvsp
Photometric mode	Abs
Measurement range	366 - 356 nm
Data interval	0.1 nm
Bandwidth	1.0 nm
Response	0.06 sec
Scan mode	Continuous
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline


Creation	QPack Soluções Analíticas Ltda
Review	
Approval	

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	360.7	1.01638			

Creation QPack Soluções Analíticas Ltda
Review
Approval

Full Width at Half Maximum - 03-Resolução.jws



[Result of FWHM Calculation]

Base 1	Base 2	Peak	Height	FWHM
355	368.4	360.7	0.914943	2.20242

[Comments]

Sample name	Repetibilidade do Comprimento de Onda_A3
Comment	Qualificação
User	QPackSA
Workgroup	Jasco
Division	Suporte
Company	ACG do Brasil

[Detailed Information]

Creation date	17/06/2026 13:32
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	370 nm
End	350 nm
Data interval	0.1 nm
Data points	201

[Measurement Information]

Instrument name	V-730
Model name	V-730
Serial No.	A149461798
Measurement date	17/06/2026 13:29
Parameter file	C:\Jasco\Qualificação\2024\Resolução_2026.uvsp
Photometric mode	Abs
Measurement range	370 - 350 nm
Data interval	0.1 nm
Bandwidth	1.0 nm
Response	0.06 sec
Scan mode	Continuous
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline
No. of cycles	1

Creation
Review
Approval

Unsigned

04-Exatidão Fotométrica

Sample name	
Comment	
User	QPackSA
Division	Suporte
Company	ACG do Brasil
Creation date	17/06/2026 14:04
Last update	17/06/2026 14:11
Instrument name	V-730
Model name	V-730
Serial No.	A149461798
ROM version	1.01.16
Parameter file	C:\Jasco\Rep Foto.uvfw
Photometric mode	Abs
Bandwidth	1.0 nm
Response	0.96 sec
Wavelength-1	440.0 nm
Change source at	340 nm
Light source	D2/WI
Dark correction	Off
Blank correction	On/Measured
No. of cycles	1

Mode	Sample Name	Comment	440.0 nm
Blank-1			-0.0385
Blank-2			-0.0386
Sample-1			0.3232
Sample-2			0.5001
Sample-3			1.0658

Unsigned

Creation
Review
Approval

Sample name	
Comment	
User	QPackSA
Division	Suporte
Company	ACG do Brasil
Creation date	17/06/2026 14:21
Last update	17/06/2026 14:22
Instrument name	V-730
Model name	V-730
Serial No.	A149461798
ROM version	1.01.16
Parameter file	C:\Jasco\Qualificação\2026\Repetibilidade Fotométrica.uvfw
Photometric mode	Abs
Bandwidth	1.0 nm
Response	0.96 sec
Wavelength-1	361.0 nm
Change source at	340 nm
Light source	D2/WI
Dark correction	Off
Blank correction	On/Measured
No. of cycles	3
Cycle interval	0 sec

Mode	Sample Name	Comment	361.0 nm
Blank-1			-0.0166
ample-1-			0.9732
ample-1-			0.9736
ample-1-			0.9733
mple-1-A			0.9733
mple-1-S			0.0002
mple-1-C			0.0200

Unsigned

Creation
Review
Approval


Sample name	
Comment	
User	QPackSA
Division	Suporte
Company	ACG do Brasil
Creation date	17/06/2026 14:24
Last update	17/06/2026 14:26
Instrument name	V-730
Model name	V-730
Serial No.	A149461798
ROM version	1.01.16
Parameter file	C:\Jasco\Linearidade Fotométrica.uvfw
Photometric mode	Abs
Bandwidth	1.0 nm
Response	0.96 sec
Wavelength-1	440.0 nm
Change source at	340 nm
Light source	D2/WI
Dark correction	Off
Blank correction	On/Measured
No. of cycles	1

Mode	Sample Name	Comment	440.0 nm
Blank-1			-0.0386
Sample-1			0.3234
Sample-2			1.0662
Sample-3			0.5002

Unsigned

Creation
Review
Approval

Peak Find - 07-Estabilidade da Linha de Base.jws



[Comments]

Sample name	Estabilidade da linha de Base
Comment	
User	QPackSA
Workgroup	Jasco
Division	Suporte
Company	ACG do Brasil

[Detailed Information]

Creation date	17/06/2026 14:50
Data array type	Linear data array
Horizontal axis	Time [sec]
Vertical axis	Abs
Start	0 sec
End	900 sec
Data interval	1 sec
Data points	901

[Measurement Information]

Instrument name	V-730
Model name	V-730
Serial No.	A149461798
Measurement date	17/06/2026 14:44
Parameter file	C:\Jasco\Qualificação\2024\07 - Estabilidade da linha de Base.uvtm
Photometric mode	Abs
Measurement range	0 - 900 sec
Data interval	1 sec
UV/Vis bandwidth	1.0 nm
Response	0.96 sec
Monitor wavelength	250 nm
Change source at	340 nm
Light source	D2/WI
Correction	Blank


[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	781	5.0549E-005	2	791	-0.000115042

Creation
Review
Approval

Unsigned

Peak Find - 07-Estabilidade da Linha de Base.jws



[Comments]

Sample name	Estabilidade da linha de Base
Comment	
User	QPackSA
Workgroup	Jasco
Division	Suporte
Company	ACG do Brasil

[Detailed Information]

Creation date	17/06/2026 14:50
Data array type	Linear data array
Horizontal axis	Time [sec]
Vertical axis	Abs
Start	0 sec
End	900 sec
Data interval	1 sec
Data points	901

[Measurement Information]


Instrument name	V-730
Model name	V-730
Serial No.	A149461798
Measurement date	17/06/2026 14:44
Parameter file	C:\Jasco\Qualificação\2024\07 - Estabilidade da linha de Base.uvtm
Photometric mode	Abs
Measurement range	0 - 900 sec
Data interval	1 sec
UV/Vis bandwidth	1.0 nm
Response	0.96 sec
Monitor wavelength	250 nm
Change source at	340 nm
Light source	D2/WI
Correction	Blank

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	412	6.97705E-005	2	415	-3.2963E-005

Creation	QPack Soluções Analíticas Ltda
Review	
Approval	

Peak Find - 08-Uniformidade da Linha de Base.jws



[Comments]

Sample name	Repetibilidade do Comprimento de Onda_A3
Comment	Qualificação
User	QPackSA
Workgroup	Jasco
Division	Suporte
Company	ACG do Brasil

[Detailed Information]

Creation date	17/06/2026 15:03
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	1000 nm
End	200 nm
Data interval	1 nm
Data points	801

[Measurement Information]

Instrument name	V-730
Model name	V-730
Serial No.	A149461798
Measurement date	17/06/2026 14:59
Parameter file	C:\Jasco\Qualificação\2026\Uniformidade Da Linha De Base_2026.uvsp
Photometric mode	Abs
Measurement range	1000 - 200 nm
Data interval	1 nm
Bandwidth	1.0 nm
Response	0.06 sec
Scan mode	Continuous
Scan speed	400 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline

Creation	QPack Soluções Analíticas Ltda
Review	
Approval	

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	406	0.00122558	2	347	-0.00136932

Sample name	
Comment	
User	QPackSA
Division	Suporte
Company	ACG do Brasil
Creation date	17/06/2026 15:27
Last update	17/06/2026 15:27
Instrument name	V-730
Model name	V-730
Serial No.	A149461798
ROM version	1.01.16
Parameter file	C:\Jasco\Qualificação\2024\Stray Light_2025.uvfw
Photometric mode	%T
Bandwidth	1.0 nm
Response	0.24 sec
Wavelength-1	360.0 nm
Change source at	340 nm
Light source	D2/WI
Dark correction	Off
Blank correction	On/Measured
No. of cycles	1

Mode	Sample Name	Comment	360.0 nm
Blank-1			103.8937
Blank-2			99.9696
Sample-1			0.0173

Unsigned

Creation
Review
Approval