
Peak Find - Exatidão Do Comprimento De Onda.jws



[Comments]
Sample name
Comment

User	QPackSA
Workgroup	QPack_Jasco
Division	Jasco do Brasil
Company	BLANVER

[Detailed Information]

Creation date	11/07/2024 11:55
Date modified	22/07/2024 12:38
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	650 nm
End	230 nm
Data interval	0.1 nm
Data points	4201

[Measurement Information]

Instrument name	V-630
Model name	V-630
Serial No.	B149061148
Accessory	USE-753
Accessory S/N	B149061148
Cell length	10 mm
Measurement date	11/07/2024 11:53
Parameter file	C:\QPack\Exatidão Do Comprimento De Onda.uvsp
Photometric mode	Abs
Measurement range	650 - 230 nm
Data interval	0.1 nm
	1.5 nm
Response	Fast
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	None

Creation
Review
Approval

Unsigned


[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity	No.	Position	Intensity
1	637.5	0.201844	2	536.4	0.409084	3	360.9	0.981148
4	279.8	0.695629						

Unsigned

Creation
Review
Approval

Peak Find - Memory-1



[Comments]
Sample name
Comment

User	QPackSA
Workgroup	QPack_Jasco
Division	Jasco do Brasil
Company	BLANVER

[Detailed Information]

Creation date	22/07/2024 13:03
Date modified	22/07/2024 13:05
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	366 nm
End	350 nm
Data interval	1 nm
Data points	17

[Measurement Information]

Instrument name	V-630
Model name	V-630
Serial No.	B149061148
Accessory	USE-753
Accessory S/N	B149061148
Cell length	10 mm
Measurement date	22/07/2024 13:03
Parameter file	C:\QPack\Repetibilidade de Comp. Onda.uvsp
Photometric mode	Abs
Measurement range	366 - 350 nm
Data interval	1 nm
	1.5 nm
Response	Fast
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline

Creation
Review
Approval

Unsigned


[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	361	0.619241			

Unsigned

Creation
Review
Approval

Peak Find - Memory-1



[Comments]
Sample name
Comment

User	QPackSA
Workgroup	QPack_Jasco
Division	Jasco do Brasil
Company	BLANVER

[Detailed Information]

Creation date	22/07/2024 13:08
Date modified	22/07/2024 13:09
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	366 nm
End	350 nm
Data interval	1 nm
Data points	17

[Measurement Information]

Instrument name	V-630
Model name	V-630
Serial No.	B149061148
Accessory	USE-753
Accessory S/N	B149061148
Cell length	10 mm
Measurement date	22/07/2024 13:08
Parameter file	C:\QPack\Repetibilidade de Comp. Onda.uvsp
Photometric mode	Abs
Measurement range	366 - 350 nm
Data interval	1 nm
	1.5 nm
Response	Fast
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline

Creation
Review
Approval

Unsigned


[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	361	0.623039			

Unsigned

Creation
Review
Approval

Peak Find - Memory-1



[Comments]
Sample name
Comment

User	QPackSA
Workgroup	QPack_Jasco
Division	Jasco do Brasil
Company	BLANVER

[Detailed Information]

Creation date	22/07/2024 13:12
Date modified	22/07/2024 13:13
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	366 nm
End	350 nm
Data interval	1 nm
Data points	17

[Measurement Information]

Instrument name	V-630
Model name	V-630
Serial No.	B149061148
Accessory	USE-753
Accessory S/N	B149061148
Cell length	10 mm
Measurement date	22/07/2024 13:12
Parameter file	C:\QPack\Repetibilidade de Comp. Onda.uvsp
Photometric mode	Abs
Measurement range	366 - 350 nm
Data interval	1 nm
	1.5 nm
Response	Fast
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline

Creation
Review
Approval


Unsigned

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	361	0.624701			

Unsigned

Creation
Review
Approval



[Comments]
Sample name
Comment

User	QPackSA
Workgroup	QPack_Jasco
Division	Jasco do Brasil
Company	BLANVER

Memory-1

[Detailed Information]

Creation date	22/07/2024 13:18
Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	370 nm
End	350 nm
Data interval	1 nm
Data points	21

[Measurement Information]

Instrument name	V-630
Model name	V-630
Serial No.	B149061148
Accessory	USE-753
Accessory S/N	B149061148
Cell length	10 mm
Measurement date	22/07/2024 13:18
Parameter file	C:\QPack\Resoluçao.uvsp
Photometric mode	Abs
Measurement range	370 - 350 nm
Data interval	1 nm
	1.5 nm
Response	Fast
Scan speed	40 nm/min
Change source at	340 nm
Light source	D2/WI
Filter exchange	Step
Correction	Baseline

Creation
Review
Approval

Unsigned

Sem título

Sample name	
Comment	
User	QPackSA
Division	Jasco do Brasil
Company	BLANVER
Creation date	22/07/2024 14:27
Last update	22/07/2024 14:35
Instrument name	V-630
Model name	V-630
Serial No.	B149061148
ROM version	1.19.00
Accessory	USE-753
Accessory S/N	B149061148
Cell length	10 mm
Parameter file	C:\QPack\Exatidão Fotométrica PN 440 fin.uvfw
Photometric mode	Abs
	1.5 nm
Response	Medium
Wavelength-1	440.0 nm
Change source at	340 nm
Light source	D2/WI
Dark correction	Off

No.	Mode	Sample Name	Comment	440.0 nm
1	Sample-1	F3 - A588/3		0.5160
2	Sample-2	F4 - A588/4		0.9711
3	Sample-3	F6 - A558/6		2.0551

Unsigned


Creation
Review
Approval

Sample name	
Comment	
User	QPackSA
Division	Jasco do Brasil
Company	BLANVER
Creation date	22/07/2024 14:48
Last update	22/07/2024 14:49
Instrument name	V-630
Model name	V-630
Serial No.	B149061148
ROM version	1.19.00
Accessory	USE-753
Accessory S/N	B149061148
Cell length	10 mm
Parameter file	C:\QPack\Repetibilidade Fotometrica -.uvfw
Photometric mode	Abs
	1.5 nm
Response	Medium
Wavelength-1	361.0 nm
Change source at	340 nm
Light source	D2/WI
Dark correction	Off
No. of cycles	3
Cycle interval	0 sec

No.	Mode	Sample Name	Comment	361.0 nm
1	Sample-1-			0.6186
2	Sample-1-			0.6182
3	Sample-1-			0.6186
4	mple-1-Av			0.6184
5	mple-1-S.			0.0002
6	mple-1-C			0.0384

Unsigned

Creation
Review
Approval



Memory-1

[Comments]
Sample name
Comment
User QPackSA
Workgroup QPack_Jasco
Division Jasco do Brasil
Company BLANVER

[Detailed Information]
Creation date 22/07/2024 15:27

Data array type Linear data array
Horizontal axis Time [sec]
Vertical axis Abs
Start 0 sec
End 900 sec
Data interval 5 sec
Data points 181

[Measurement Information]
Instrument name V-630
Model name V-630
Serial No. B149061148

Accessory USE-753
Accessory S/N B149061148
Cell length 10 mm

Measurement date 22/07/2024 15:26


Parameter file C:\QPack\Estab. Linha Base.uvtm
Photometric mode Abs
Measurement range 0 - 900 sec
Data interval 5 sec
UV/Vis bandwidth 1.5 nm
Response Slow
Monitor wavelength 250 nm
Change source at 340 nm
Light source D2/WI
Correction Off

Creation
Review
Approval

Unsigned

Inspection Results

Date	segunda-feira, 22 de julho de 2024
Model name	V-630
Serial No.	B149061148



Baseline Flatness	Criteria: <= 0.00500 Abs Minimum: -0.00008 Abs, Maximum: 0.00647 Abs, (Max.-Min.)/2: 0.00327 Abs, Pass

Unsigned

Creation
Review
Approval

Inspection Results

Date	segunda-feira, 22 de julho de 2024
Model name	V-630
Serial No.	B149061148

Stray Light Pass/Fail : Pass

Item	Measured	Criteria	Pass/Fail
370.0 nm	NaNO2	0.88237 %	<= 1.00000

Unsigned

Creation
Review
Approval