
Peak Find - EDUV.jws



[Comments]


User	Mateus
Division	
Company	QPack
Creation date	05/09/2024 16:13

[Detailed Information]		[Measurement Information]
Data array type	Linear data array	Model Name
Horizontal axis	Wavelength [nm]	Serial No.
Vertical axis	Abs	
Start	200 nm	
End	600 nm	
Data interval	2 nm	
Data points	200	

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	272	0.106607	2	244	0.031185
3	204	0.280693			

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 14:23:21
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 14:20:18
Volume	10.0 [µL]
Sample #	1
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 01-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 14:28:30
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 14:25:26
Volume	10.0 [µL]
Sample #	2
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 01-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,050	711926	98014	100,000	100,000	N/A	541	N/A	1,912		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 14:33:33
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 14:30:31
Volume	10.0 [µL]
Sample #	2
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 02-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1.058	710003	98124	100.000	100.000	N/A	557	N/A	1.855		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 14:38:41
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 14:35:39
Volume	10.0 [µL]
Sample #	2
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 03-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,058	708883	96966	100,000	100,000	N/A	545	N/A	1,969		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 14:43:49
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 14:40:47
Volume	10.0 [µL]
Sample #	1
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 02-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 14:48:55
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 14:45:52
Volume	10.0 [µL]
Sample #	2
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 04-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1.067	708697	97253	100.000	100.000	N/A	556	N/A	1.894		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 14:54:03
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 14:51:00
Volume	10.0 [µL]
Sample #	2
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 05-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,075	707843	97263	100,000	100,000	N/A	565	N/A	1,794		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 14:59:11
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 14:56:09
Volume	10.0 [µL]
Sample #	2
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 06-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,075	707241	96881	100,000	100,000	N/A	562	N/A	1,894		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 15:04:19
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:01:16
Volume	10.0 [µL]
Sample #	1
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 03-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 15:09:27
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:06:24
Volume	10.0 [µL]
Sample #	3
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	LDUV 01-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 15:14:35
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:11:32
Volume	10.0 [µL]
Sample #	4
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 02-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,075	239847	32798	100,000	100,000	N/A	555	N/A	1,750		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 15:19:43
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:16:40
Volume	10.0 [µL]
Sample #	5
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 03-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,075	452514	61964	100,000	100,000	N/A	556	N/A	1,898		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 15:24:53
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:21:51
Volume	10.0 [µL]
Sample #	6
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 04-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,075	708177	96717	100,000	100,000	N/A	553	N/A	1,764		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 15:30:03
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:27:01
Volume	10.0 [µL]
Sample #	7
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 05-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1.067	927766	127300	100.000	100.000	N/A	547	N/A	1.768		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 15:52:21
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:32:19
Volume	20,0 [µL]
Sample #	8
Project Name	Qualificação
Acquisition Time	20,0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	RELB HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	RELB 01-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,083	733122	110147	100,000	100,000	N/A	709	N/A	1,726		11508,4

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 15:52:21
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:32:19
Volume	20,0 [µL]
Sample #	8
Project Name	Qualificação
Acquisition Time	20,0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	RELB HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	RELB 01-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,083	733122	110147	100,000	100,000	N/A	709	N/A	1,726		11508,4

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 15:52:21
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:32:19
Volume	20,0 [µL]
Sample #	8
Project Name	Qualificação
Acquisition Time	20,0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	RELB HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	RELB 01-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,083	733122	110147	100,000	100,000	N/A	709	N/A	1,726		11508,4

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	30/08/2024 08:32:41
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:54:27
Volume	0,1 [µL]
Sample #	9
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 01-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação (Manual)

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,067	8942	1054	100,000	100,000	N/A	365	N/A	1,783		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 16:02:39
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 15:59:37
Volume	5.0 [µL]
Sample #	9
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 02-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1.067	345995	44223	100.000	100.000	N/A	458	N/A	1.854		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 16:07:53
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 16:04:50
Volume	10.0 [µL]
Sample #	9
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 03-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,075	712858	96325	100,000	100,000	N/A	536	N/A	1,900		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 16:13:07
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 16:10:05
Volume	15.0 [µL]
Sample #	9
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 04-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1,083	1081145	154745	100,000	100,000	N/A	626	N/A	1,846		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	29/08/2024 16:18:25
Description	
HPLC System Name	HPLC-0004
Injection Date	29/08/2024 16:15:23
Volume	20.0 [µL]
Sample #	9
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	Qualificação HP04
Control Method	Qualificação HP04
Peak ID Table	Qualificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 05-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	Qualificação

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	Cafeina	1	1.083	1452225	218468	100.000	100.000	N/A	706	N/A	1.857		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	30/08/2024 09:35:53
Description	
HPLC System Name	HPLC-0004
Injection Date	30/08/2024 09:05:52
Volume	0,0 [µL]
Sample #	-
Project Name	Qualificação
Acquisition Time	20,0 [min]
Acquisition Sequence	
Control Method	Grad AB
Peak ID Table	Gradiente
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	Gradiente AB_001-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	(Manual)

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	20 %	1	3,608	14666590	98063	5,491	6,603	N/A	N/A	N/A	N/A		N/A
2	40 %	1	6,558	32522469	198295	12,175	13,353	N/A	N/A	N/A	N/A		N/A
3	60 %	1	9,600	51648621	298485	19,336	20,099	N/A	N/A	N/A	N/A		N/A
4	80 %	1	12,617	68965103	397501	25,818	26,767	N/A	N/A	N/A	N/A		N/A
5	100 %	1	15,600	99313137	492723	37,180	33,178	N/A	N/A	N/A	N/A		N/A

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	30/08/2024 12:03:37
Description	
HPLC System Name	HPLC-0004
Injection Date	30/08/2024 09:28:36
Volume	0,0 [µL]
Sample #	-
Project Name	Qualificação
Acquisition Time	20,0 [min]
Acquisition Sequence	
Control Method	Grad CD
Peak ID Table	Gradiente
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Gradiente CD_002-CH1
Sample Name	
Channel Name	UV
Sampling Interval	500 [msec]
Peak Method	(Manual)

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning	S/N
1	20 %	1	3,458	14098096	100260	5,249	6,740	N/A	N/A	N/A	N/A		N/A
2	40 %	1	6,558	34415700	199383	12,813	13,403	N/A	N/A	N/A	N/A		N/A
3	60 %	1	9,575	51419410	298091	19,144	20,039	N/A	N/A	N/A	N/A		N/A
4	80 %	1	12,525	68882988	397413	25,646	26,716	N/A	N/A	N/A	N/A		N/A
5	100 %	1	15,617	99778355	492419	37,148	33,102	N/A	N/A	N/A	N/A		N/A

Creator QPackSA
Reviewer
Approver