
Peak Find - Comprimento de Onda.jws



[Comments]

User	QPackSA
Division	---
Company	DLA PHARMA
Creation date	19/10/2022 08:37
Date modified	19/10/2022 08:38

[Detailed Information]

Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	200 nm
End	300 nm
Data interval	1 nm
Data points	101


[Measurement Information]

Instrument Name	X-LC 3110MD
Model Name	X-LC 3110MD
Serial No.	A002661329
ROM version	V1.05.01

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	272	0.1277	2	244	0.036526
3	207	0.302417			

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 08:55:51
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 08:52:49
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 PCIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 01-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 08:59:31
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 08:56:29
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 PCIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 01-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,181	750542	131726	100,000	100,000	N/A	3534	N/A	1,236	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:03:11
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:00:09
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 PCIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 02-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,178	752923	132523	100,000	100,000	N/A	3556	N/A	1,243	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:06:51
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:03:49
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 PCIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 03-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,158	748194	132831	100,000	100,000	N/A	3539	N/A	1,274	

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:10:33
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:07:31
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 PCIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 02-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:14:13
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:11:11
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 PCIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 04-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,141	753225	134474	100,000	100,000	N/A	3546	N/A	1,256	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:17:53
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:14:52
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 PCIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 05-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,136	749370	134584	100,000	100,000	N/A	3570	N/A	1,260	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:21:33
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:18:32
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 PCIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN 06-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,129	748456	135253	100,000	100,000	N/A	3589	N/A	1,286	

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:25:13
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:22:12
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 PCIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 03-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:32:47
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:29:37
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LDUV
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	LDUV 01-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:36:19
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:33:17
Volume	10,0 [µL]
Sample #	2
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LDUV
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 02-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,115	245245	44346	100,000	100,000	N/A	3566	N/A	1,278	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:40:03
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:37:00
Volume	10,0 [µL]
Sample #	3
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LDUV
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 03-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,121	503920	92109	100,000	100,000	N/A	3641	N/A	1,263	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:43:45
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:40:42
Volume	10,0 [µL]
Sample #	4
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LDUV
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 04-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,118	743824	136451	100,000	100,000	N/A	3638	N/A	1,268	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 09:47:27
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 09:44:25
Volume	10,0 [µL]
Sample #	5
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LDUV
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV 05-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,118	1000698	182900	100,000	100,000	N/A	3654	N/A	1,275	

Creator QPackSA
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 10:05:05
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 10:02:02
Volume	0,1 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LIIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	LIIN 01-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 10:08:41
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 10:05:39
Volume	5,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LIIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 02-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,098	363521	67222	100,000	100,000	N/A	3623	N/A	1,269	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 10:12:25
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 10:09:22
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LIIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 03-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,094	743111	137289	100,000	100,000	N/A	3639	N/A	1,281	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 10:16:09
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 10:13:07
Volume	15,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LIIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 04-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,121	1118882	208460	100,000	100,000	N/A	3789	N/A	1,287	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 10:19:59
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 10:16:57
Volume	20,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	3,0 [min]
Acquisition Sequence	Qualificação X-LC 2022 LIIN
Control Method	QUALIFICAÇÃO
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN 05-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	50 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	2,126	1496211	276808	100,000	100,000	N/A	3767	N/A	1,289	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information

User Name	QPackSA
Date Modified	19/10/2022 12:01:01
Description	
HPLC System Name	X-LC
Injection Date	19/10/2022 11:23:18
Volume	0,0 [µL]
Sample #	–
Project Name	Qualificação QPack
Acquisition Time	20,0 [min]
Acquisition Sequence	
Control Method	QUALIFICAÇÃO GRBO A-B
Peak ID Table	Identificação GRBO
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	Gradiente A-B HP03 2022_002-CH5
Sample Name	
Channel Name	273,0nm
Sampling Interval	100 [msec]
Peak Method	(Manual)

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	20%	5	3,413	14031969	84882	6,264	6,831	N/A	N/A	N/A	N/A	
2	40%	5	6,868	29340415	168868	13,098	13,590	N/A	N/A	N/A	N/A	
3	60%	5	9,433	43788550	250547	19,548	20,164	N/A	N/A	N/A	N/A	
4	80%	5	12,230	59030040	329932	26,352	26,552	N/A	N/A	N/A	N/A	
5	100%	5	14,818	77814561	408343	34,738	32,863	N/A	N/A	N/A	N/A	

Creator QPackSA
Reviewer
Approver

Peak Find - Comprimento de Onda Detector FL



[Comments]

User	QPackSA
Division	---
Company	DLA PHARMA
Creation date	19/10/2022 13:55

[Detailed Information]

Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Int.
Start	360 nm
End	700 nm
Data interval	2 nm
Data points	171


[Measurement Information]

Instrument Name	X-LC 3120FP
Model Name	X-LC 3120FP
Serial No.	007961214
ROM version	V6.34
Measurement mode	Emission
Ex Wavelength	350 nm

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	450	0.665728			

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 14:46:36
Description	
HPLC System Name	X-LC Det FL
Injection Date	19/10/2022 14:19:28
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	2,0 [min]
Acquisition Sequence	QUALIFICAÇÃO HP03 – PDFL
Control Method	Qualificação HP03 2022
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PDFL_01-CH1
Sample Name	
Channel Name	CH1
Sampling Interval	200 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	PD Fluorescência	1	0,117	91376	26928	100,000	100,000	N/A	26	N/A	1,629	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 14:46:36
Description	
HPLC System Name	X-LC Det FL
Injection Date	19/10/2022 14:23:58
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	2,0 [min]
Acquisition Sequence	QUALIFICAÇÃO HP03 – PDFL
Control Method	Qualificação HP03 2022
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PDFL_02-CH1
Sample Name	
Channel Name	CH1
Sampling Interval	200 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	PD Fluorescência	1	0,117	91063	26730	100,000	100,000	N/A	25	N/A	1,580	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 14:46:36
Description	
HPLC System Name	X-LC Det FL
Injection Date	19/10/2022 14:28:27
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	2,0 [min]
Acquisition Sequence	QUALIFICAÇÃO HP03 – PDFL
Control Method	Qualificação HP03 2022
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PDFL_03-CH1
Sample Name	
Channel Name	CH1
Sampling Interval	200 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	PD Fluorescência	1	0,117	90940	26737	100,000	100,000	N/A	25	N/A	1,650	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 14:46:36
Description	
HPLC System Name	X-LC Det FL
Injection Date	19/10/2022 14:32:57
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	2,0 [min]
Acquisition Sequence	QUALIFICAÇÃO HP03 – PDFL
Control Method	Qualificação HP03 2022
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PDFL_04-CH1
Sample Name	
Channel Name	CH1
Sampling Interval	200 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	PD Fluorescência	1	0,117	90522	26612	100,000	100,000	N/A	25	N/A	1,571	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 14:46:36
Description	
HPLC System Name	X-LC Det FL
Injection Date	19/10/2022 14:37:26
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	2,0 [min]
Acquisition Sequence	QUALIFICAÇÃO HP03 – PDFL
Control Method	Qualificação HP03 2022
Peak ID Table	Identificação
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PDFL_05-CH1
Sample Name	
Channel Name	CH1
Sampling Interval	200 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	PD Fluorescência	1	0,117	91655	27130	100,000	100,000	N/A	26	N/A	1,623	

Creator QPackSA
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	QPackSA
Date Modified	19/10/2022 14:46:37
Description	
HPLC System Name	X-LC Det FL
Injection Date	19/10/2022 14:41:56
Volume	10,0 [µL]
Sample #	1
Project Name	Qualificação QPack
Acquisition Time	2,0 [min]
Acquisition Sequence	QUALIFICAÇÃO HP03 – PDFL
Control Method	Qualificação HP03 2022
Peak ID Table	Identificação
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	PDFL_06-CH1
Sample Name	
Channel Name	CH1
Sampling Interval	200 [msec]
Peak Method	Integração

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	PD Fluorescência	1	0,117	90860	26448	100,000	100,000	N/A	25	N/A	1,539	

Creator QPackSA
Reviewer
Approver