Single Injection Report

Agilent Technologies


Data file:	C:\CDSProjects\Qualificação\Results\EDUV 001.sirslt\EDUV 001.dx		
Sample name:	EDUV 001		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A1
Injection date:	2025-08-28 14:48:57-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação - Spectro.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 12:17:23-03:00		

TR:

1.751

Nome do Composto:

Cafeína




Name	RT [min]	Area	Peak Height
Cafeína	1.751	291.3428	57.661

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\RELB 001.sirslt\RELB 001.dx		
Sample name:	RELB 001		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A2
Injection date:	2025-08-28 14:52:57-03:00	**Injection:**	1 of 1
Acq. method:	Sinal-Ruido.amx	**Injection volume:**	20.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.820	789.695	142.452

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\RELB 001.sirslt\RELB 001.dx		
Sample name:	RELB 001		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A2
Injection date:	2025-08-28 14:52:57-03:00	**Injection:**	1 of 1
Acq. method:	Sinal-Ruido.amx	**Injection volume:**	20.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.820	789.695	142.452

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\RELB 001.sirslt\RELB 001.dx		
Sample name:	RELB 001		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A2
Injection date:	2025-08-28 14:52:57-03:00	**Injection:**	1 of 1
Acq. method:	Sinal-Ruido.amx	**Injection volume:**	20.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.820	789.695	142.452

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\RELB 001.sirslt\RELB 001.dx		
Sample name:	RELB 001		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A2
Injection date:	2025-08-28 14:52:57-03:00	**Injection:**	1 of 1
Acq. method:	Sinal-Ruido.amx	**Injection volume:**	20.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.820	789.695	142.452

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\PCIN 001.rslt\Branco PCIN 01.dx		
Sample name:	Branco PCIN 01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A3
Injection date:	2025-08-28 15:13:52-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.828	2.184	0.441

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\PCIN 001.rslt\PCIN 01.dx		
Sample name:	PCIN 01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A4
Injection date:	2025-08-28 15:17:37-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.833	837.241	144.418

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\PCIN 001.rslt\PCIN 02.dx		
Sample name:	PCIN 02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A4
Injection date:	2025-08-28 15:21:24-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.827	840.209	145.779

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\PCIN 001.rslt\PCIN 03.dx		
Sample name:	PCIN 03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A4
Injection date:	2025-08-28 15:25:10-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.825	838.895	145.761

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\PCIN 001.rslt\Branco PCIN 02.dx		
Sample name:	Branco PCIN 02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A3
Injection date:	2025-08-28 15:28:55-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.829	2.626	0.518

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\PCIN 001.rslt\PCIN 04.dx		
Sample name:	PCIN 04		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A4
Injection date:	2025-08-28 15:32:41-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.822	841.002	146.463

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\PCIN 001.rslt\PCIN 05.dx		
Sample name:	PCIN 05		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A4
Injection date:	2025-08-28 15:36:27-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.824	838.560	147.349

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\PCIN 001.rslt\PCIN 06.dx		
Sample name:	PCIN 06		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A4
Injection date:	2025-08-28 15:40:12-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.825	839.818	147.860

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\PCIN 001.rslt\Branco PCIN 03.dx		
Sample name:	Branco PCIN 03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A3
Injection date:	2025-08-28 15:43:59-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-27 14:11:23-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.824	3.290	0.604

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LDUV 002.rslt\LDUV 01.dx		
Sample name:	LDUV 01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A5
Injection date:	2025-08-28 16:13:14-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:33:12-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.674	6.837	0.517

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LDUV 002.rslt\LDUV 02.dx		
Sample name:	LDUV 02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A6
Injection date:	2025-08-28 16:16:59-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:33:12-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.825	266.686	47.499

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LDUV 002.rslt\LDUV 03.dx		
Sample name:	LDUV 03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A7
Injection date:	2025-08-28 16:20:43-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:33:12-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.823	526.403	95.561

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LDUV 002.rslt\LDUV 04.dx		
Sample name:	LDUV 04		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A8
Injection date:	2025-08-28 16:24:29-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:33:12-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.825	785.133	144.824

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LDUV 002.rslt\LDUV 05.dx		
Sample name:	LDUV 05		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A9
Injection date:	2025-08-28 16:28:15-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:33:12-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.826	1063.972	192.332

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LIIN 002.rslt\LIIN 01.dx		
Sample name:	LIIN 01		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A10
Injection date:	2025-08-28 16:32:07-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	0.100
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:51:26-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.815	8.735	1.455

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LIIN 002.rslt\LIIN 02.dx		
Sample name:	LIIN 02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A10
Injection date:	2025-08-28 16:35:49-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	5.000
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:51:26-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.816	376.186	70.006

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LIIN 002.rslt\LIIN 03.dx		
Sample name:	LIIN 03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A10
Injection date:	2025-08-28 16:39:35-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	10.000
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:51:26-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.826	745.662	140.228

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LIIN 002.rslt\LIIN 04.dx		
Sample name:	LIIN 04		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A10
Injection date:	2025-08-28 16:43:23-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	15.000
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:51:26-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.831	1122.129	210.155

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\LIIN 002.rslt\LIIN 05.dx		
Sample name:	LIIN 05		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	P1-A10
Injection date:	2025-08-28 16:47:15-03:00	**Injection:**	1 of 1
Acq. method:	Qualificação.amx	**Injection volume:**	20.000
Analysis method:	*Qualificação.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-28 16:51:26-03:00		




Name	RT [min]	Area	Peak Height
Cafeina	1.836	1510.462	283.557

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\GRAD 001.rslt\GRAD AB 02.dx		
Sample name:	GRAD AB 02		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	
Injection date:	2025-08-28 17:27:36-03:00	**Injection:**	1 of 1
Acq. method:	Grad AB.amx	**Injection volume:**	0.000
Analysis method:	*Grad AB.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-29 09:03:12-03:00		




Name	RT [min]	Area	Peak Height
20%	4.165	1139.769	37.706
40%	6.852	12095.196	76.582
60%	9.758	19547.341	115.751
80%	12.663	28594.188	154.560
100%	16.305	38280.325	195.867

Single Injection Report

Agilent Technologies

Data file:	C:\CDSProjects\Qualificação\Results\2025\CL-LABMAR-004\GRAD 001.rslt\GRAD CD 03.dx		
Sample name:	GRAD CD 03		
Description:			
Sample amount:	0.000	**Sample type:**	Sample
Instrument:	CL-LABMAR-004	**Location:**	
Injection date:	2025-08-28 18:48:04-03:00	**Injection:**	1 of 1
Acq. method:	Grad CD.amx	**Injection volume:**	0.000
Analysis method:	*Grad AB.pmx	**Acq. operator:**	QPackSA
Last changed:	2025-08-29 09:07:56-03:00		



Name	RT [min]	Area	Peak Height
20%	4.079	764.473	38.765
40%	6.347	9352.286	77.613
60%	9.458	15465.952	116.639
80%	13.027	26923.800	155.440
100%	16.183	39569.292	195.724