Resultado obtido na altura do pico

	ALTURA DO PICO (mAU)
Resultado:	**105,0000**


23/09/2026 09:42:11 Page 1 / 1

SHIMADZU LabSolutions **Analysis Report**

<Sample Information>

Sample Name	: RELB		
Sample ID	:		
Data Filename	: RELB.lcd		
Method Filename	: OQ_RELB.lcm		
Batch Filename	: RELB_SEQ.lcb		
Vial #	: 30	Sample Type	: Unknown
Injection Volume	: 20 uL		
Date Acquired	: 22/09/2026 16:33:23	Acquired by	: System Administrator
Date Processed	: 22/09/2026 16:53:24	Processed by	: System Administrator

<Chromatogram>




<Peak Table>

Detector A 250nm

Peak#	Ret. Time	Area	Height	Conc.	Unit	Mark	Name
1	1,051	577799	101634	0,000	g/L		Cafeina
Total		577799	101634				

Resultado obtido no teste de sinal/ruído

Resultado:	Ruído (mAU)
VALOR MÁXIMO:	0,0415
VALOR MÍNIMO:	0,0370
	Ruído
	0,0045



Resultado obtido no teste estabilidade da linha de base

Estabilidade da Linha de Base	Diferença 20 min	Diferença 60 min
VALOR MÁXIMO:	0,0800	0,2400
VALOR MÍNIMO:	0,0000	
Resultado:	0,0800	

