
Peak Find - Exploratória_001 PDA_1.8600.JWS



[Comments]

User	Administrator
Division	---
Company	AS ERVAS CURAM
Creation date	15/02/2021 16:07

[Detailed Information]

Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	200 nm
End	300 nm
Data interval	1 nm
Data points	101


[Measurement Information]

Instrument Name	MD-2010
Model Name	MD-2010
Serial No.	0000000000
ROM version	V2.36

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	273	0.027111	2	245	0.006972
3	206	0.074859			

Chromatogram




Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 16:27:40
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 16:24:37
Volume	10.00 [µL]
Sample #	11
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	PCIN
Control Method	Qualificação
Peak ID Table	ID
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 01-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 16:39:10
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 16:29:33
Volume	10.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	PCIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN01-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.827	715127	82050	100.000	100.000	N/A	1058	N/A	1.326	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 16:38:33
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 16:34:29
Volume	10.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	PCIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN02-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.820	716441	82532	100.000	100.000	N/A	1051	N/A	1.336	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 16:45:38
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 16:39:25
Volume	10.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	PCIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN03-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.820	715324	82604	100.000	100.000	N/A	1055	N/A	1.331	

Creator Administrator
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 16:47:20
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 16:44:19
Volume	10.00 [µL]
Sample #	11
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	PCIN
Control Method	Qualificação
Peak ID Table	ID
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 02-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 16:52:55
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 16:49:15
Volume	10.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	PCIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN04-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.813	717889	83142	100.000	100.000	N/A	1057	N/A	1.293	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 17:00:16
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 16:54:11
Volume	10.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	PCIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN05-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.813	721153	83270	100.000	100.000	N/A	1051	N/A	1.276	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 17:02:08
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 16:59:06
Volume	10.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	PCIN
Control Method	Qualificação
Peak ID Table	ID
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	PCIN06-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Unknown	5	1.807	719321	83288	100.000	100.000	N/A	1049	N/A	1.288	

Creator Administrator
Reviewer
Approver

Chromatogram




Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 17:08:11
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 17:03:58
Volume	10.00 [µL]
Sample #	11
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	PCIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Branco 03-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

Creator	Administrator
Reviewer	
Approver	

Chromatogram




Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 18:51:10
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 17:11:06
Volume	10.00 [µL]
Sample #	11
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LDUV
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LDUV
Additional Information	

Channel & Peak Information Table

Chromatogram Name	LDUV01-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 18:51:11
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 17:15:59
Volume	10.00 [µL]
Sample #	12
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LDUV
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LDUV
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV02-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.800	245685	28401	100.000	100.000	N/A	1046	N/A	1.327	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 18:51:12
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 17:20:53
Volume	10.00 [µL]
Sample #	13
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LDUV
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LDUV
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV03-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.800	498184	57817	100.000	100.000	N/A	1049	N/A	1.286	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 18:51:13
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 17:25:48
Volume	10.00 [µL]
Sample #	14
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LDUV
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LDUV
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV04-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.800	716263	82893	100.000	100.000	N/A	1046	N/A	1.295	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 18:51:13
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 17:30:44
Volume	10.00 [µL]
Sample #	15
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LDUV
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LDUV
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LDUV05-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.800	969205	113343	100.000	100.000	N/A	1054	N/A	1.278	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 19:27:05
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 18:56:49
Volume	0.10 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LIIN
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN01-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.767	6270	715	100.000	100.000	N/A	871	N/A	1.225	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 19:27:06
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 19:01:40
Volume	5.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LIIN
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN02-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.767	349420	39179	100.000	100.000	N/A	929	N/A	1.343	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 19:27:06
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 19:06:34
Volume	10.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LIIN
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN03-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.767	723536	85692	100.000	100.000	N/A	1050	N/A	1.301	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 19:27:07
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 19:11:34
Volume	15.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LIIN
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN04-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.767	1106547	136971	100.000	100.000	N/A	1157	N/A	1.323	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	15/02/2021 19:27:07
Description	
HPLC System Name	11.X.I.05
Injection Date	15/02/2021 19:16:40
Volume	20.00 [µL]
Sample #	16
Project Name	Qualificação
Acquisition Time	3.0 [min]
Acquisition Sequence	LIIN
Control Method	Qualificação
Peak ID Table	ID 01
Calibration Method	LIIN
Additional Information	


Channel & Peak Information Table

Chromatogram Name	LIIN05-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	400 [msec]
Peak Method	Int

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	Cafeína	5	1.773	1492125	192101	100.000	100.000	N/A	1271	N/A	1.282	

Creator Administrator
Reviewer
Approver

Chromatogram



Chromatogram Information

User Name	Administrator
Date Modified	16/02/2021 09:01:50
Description	
HPLC System Name	11.X.I.05
Injection Date	16/02/2021 08:43:48
Volume	0.00 [µL]
Sample #	-
Project Name	Qualificação
Acquisition Time	18.0 [min]
Acquisition Sequence	
Control Method	Gradiente AB
Peak ID Table	GRAD
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	Gradiente AB_001-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	800 [msec]
Peak Method	(Manual)

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	20%	5	3.520	13050281	87511	5.275	6.356	N/A	N/A	N/A	N/A	
2	40%	5	6.560	30185062	181077	12.200	13.152	N/A	N/A	N/A	N/A	
3	60%	5	9.547	46533957	276027	18.808	20.049	N/A	N/A	N/A	N/A	
4	80%	5	12.520	63806213	369638	25.789	26.849	N/A	N/A	N/A	N/A	
5	100%	5	15.560	93841360	462498	37.928	33.593	N/A	N/A	N/A	N/A	

Creator
Reviewer
Approver

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	16/02/2021 09:38:33
Description	
HPLC System Name	11.X.I.05
Injection Date	16/02/2021 09:20:32
Volume	0.00 [µL]
Sample #	-
Project Name	Qualificação
Acquisition Time	18.0 [min]
Acquisition Sequence	
Control Method	Gradiente CD
Peak ID Table	GRAD
Calibration Method	
Additional Information	

Channel & Peak Information Table

Chromatogram Name	Gradiente CD_001-CH5
Sample Name	
Channel Name	273.0nm
Sampling Interval	800 [msec]
Peak Method	(Manual)

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]	Area%	Height%	Quantity	NTP	Resolution	Symmetry Factor	Warning
1	20%	5	3.587	13506661	89024	5.451	6.457	N/A	N/A	N/A	N/A	
2	40%	5	6.493	29621170	182904	11.954	13.266	N/A	N/A	N/A	N/A	
3	60%	5	9.560	47349541	275420	19.109	19.976	N/A	N/A	N/A	N/A	
4	80%	5	12.507	62762414	368611	25.329	26.735	N/A	N/A	N/A	N/A	
5	100%	5	15.533	94549397	462795	38.157	33.566	N/A	N/A	N/A	N/A	

Creator
Reviewer
Approver