
Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 16:40:05
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 15:17:59
Sample #	1
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	PCIN
Control Method	Metodo

Chromatogram




Peak Information – No data.

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 16:40:34
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 15:31:49
Sample #	1
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	PCIN
Control Method	Metodo

Chromatogram




Peak Information – No data.

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 16:41:02
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 15:45:41
Sample #	1
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	PCIN
Control Method	Metodo


Chromatogram



Peak Information – No data.

Creator *Administrator*
Reviewer
Approver

HPLC-006-CFQ EDUV



[Comments]

User	Administrator
Division	---
Company	Greenpharma
Creation date	30/06/2020 13:08
Date modified	30/06/2020 13:08

[Detailed Information]

Data array type	Linear data array
Horizontal axis	Wavelength [nm]
Vertical axis	Abs
Start	200 nm
End	600 nm
Data interval	2 nm
Data points	200


[Measurement Information]

Instrument Name	UV-2075
Model Name	UV-2075
Serial No.	0000000000
ROM version	V6.45

[Result of Peak Picking]

No.	Position	Intensity	No.	Position	Intensity
1	274	0.553684	2	246	0.275412
3	206	1.54084			

Chromatogram



Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 18:55:26
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 18:22:24
Volume	0,00 [µL]
Sample #	-
Project Name	Qualificação
Acquisition Time	33,0 [min]
Acquisition Sequence	
Control Method	Gradiente ABCD 06
Peak ID Table	
Calibration Method	
Additional Information	


Channel & Peak Information Table

Chromatogram Name	Gradiente_002-CH1
Sample Name	
Channel Name	CH1
Sampling Interval	40 [msec]
Peak Method	

Creator	Administrator
Reviewer	
Approver	

Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 16:42:51
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 16:17:02
Sample #	1
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LDUV
Control Method	Metodo

Chromatogram




Peak Information – No data.

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 16:42:57
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 16:20:30
Sample #	3
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LDUV
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,625	253668	29043

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 16:43:05
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 16:23:58
Sample #	4
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LDUV
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,625	498388	56669

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 16:44:42
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 16:27:25
Sample #	5
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LDUV
Control Method	Metodo

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,625	786667	88958

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 16:43:12
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 16:30:53
Sample #	6
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LDUV
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,627	1042085	116807

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 17:20:04
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 16:50:25
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LIIN 06
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,623	11970	1432

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 17:20:12
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 16:53:50
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LIIN 06
Control Method	Metodo

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,627	246604	26261

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 17:20:19
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 16:57:17
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LIIN 06
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,633	503058	56439

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 17:20:27
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 17:00:48
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LIIN 06
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,637	778931	92552

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 17:20:35
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 17:04:23
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	LIIN 06
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,641	1022974	128848

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 16:40:11
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 15:21:26
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	PCIN
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,591	776942	92559

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 16:40:19
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 15:24:54
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	PCIN
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,593	776930	92724

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 16:40:26
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 15:28:22
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	PCIN
Control Method	Metodo

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,596	778260	92884

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 16:40:40
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 15:35:17
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	PCIN
Control Method	Metodo

Chromatogram



Peak Information


#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,602	776034	92347

Creator *Administrator*
Reviewer
Approver

Chromatogram Information

User Name	Administrator
Date Modified	30/06/2020 16:40:47
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 15:38:45
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	PCIN
Control Method	Metodo

Chromatogram




Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,603	778187	92472

Creator *Administrator*
Reviewer
Approver

Chromatogram Information	
User Name	Administrator
Date Modified	30/06/2020 16:40:55
Description	
HPLC System Name	HPLC-006-CFQ
Injection Date	30/06/2020 15:42:13
Sample #	2
Project Name	Qualificação
Acquisition Time	3,0 [min]
Acquisition Sequence	PCIN
Control Method	Metodo

Chromatogram



Peak Information

#	Peak Name	CH	tR [min]	Area [µV·sec]	Height [µV]
1	Cafeína	1	1,606	778492	92024

Creator *Administrator*
Reviewer
Approver